

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 3/2/11

Product Name	CLEARKLENS DE STANDARD VH29
Product Code	7515499
Batch Number	W029511
Production Date	01/2011
Expiry Date	01/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.885

"SIGNATURE" C. Pascal

Name Caroline Pascal
Function Name Pharmacovigilance Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

CERTIFICATE OF IRRADIATION

Number: **B**

0476271 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P016468
Date Rec'd 27/01/11
Date: 2/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B211010755	JD DE 900ML BOTTLES 25-45kGy	532	BATCH NO: W029511 CUST B/NO: 2402 ACTUAL DOSE....36.3 - 37.5....kGy
B2FLE0010999 B211010755	SAMPLES	1	CUSTOMERS BATCH/CODE NO'S AS ORDER NO P016468 DATED 27.01.11
Total		533	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED