



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 8/7/11

Product Name	CLEARKLENS DE VH29
Product Code	6069643
Batch Number	12030246
Production Date	06/2011
Expiry Date	06/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	Clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.885

"SIGNATURE"

C. P. B. O. E.

Name

Caroline P. B. O. E.

Function Name

Pharmaceutical supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

CERTIFICATE OF IRRADIATION

Number: **B**

0493844 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P017119
Date Rec'd 23/06/11
Date: 4/07/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B211070072	JD DE 900ML BOTTLES 25-45kGy	510	415 X B/No. W030246 CUST/B/No. 2578 95 X B/No. W030120 ACTUAL DOSE....35.8 - 37.2....kGy
B2FLE0010999 B211070072	SAMPLES	11	 CUSTOMER BATCH/CODE No.s AS ORDER No. P017119 DATED 22/6/11
	Despatch Text: 10PLTS		
	Total	521	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 11 Jul 11
Date Tested: 16 Jul 11
Date Completed: 30 Jul 11
Purchase Order: PO17212

Test Article: Clearklens DE Block I
Sample Code: 2578
Batch Ref: WO30246
Qty Received: 20 x 900ml.
Laboratory Reference: 0001956/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM03
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 01 Aug 11 17:20:57

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001956
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004481

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS


Date Received: 11 Jul 11
Date Tested: 14 Jul 11
Date Completed: 14 Jul 11
Purchase Order: PO17212

Test Article: Clearklens DE Block 1
Sample Code: -
Batch Ref: WO30246
Qty Received: 1 x 900mL
Laboratory Reference: 0001955/1
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	127 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	MM110/05	FM01

External Comments

Test carried out according to current Pharmacopoeias.

 18/07/11

Approval is provided by Electronic Signature. Their name is shown below.



Date: 15 Jul 11 11:26:49

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001955
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004287

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited