

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 12/12/11

Product Name	CLEARKLENS DE VH29		
Product Code	6069643		
Batch Number	WO 31005		
Production Date	1/2011		
Expiration Date	11/2013		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear colourles
				119010
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name:	C More
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 6069643	Version: 03	Date of issuing: November 28th 2011
-------------	-------------	-------------------------------------

CERTIFICATE OF IRRADIATION Number: B

0512110 00

FLECIO1

Flexible Medical Fackaging Ltd

Unit 8

White Cross Industrial Estate

Hightown Lancaster

LA1 4XS

Cust. Ref: P017811
Date Rec'd 24/11/11
Date: 2/12/11



Isotron Limited

Page 1

Roydsdale Way · Euroway Trading Estate

Bradford · West Yorkshire BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011 Facsimile: +44 (0)1274 686061 Email: assistance@isotron.com

Website: www.isotron.com

Date: 2/12/	(11		
ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
82FLE0010121 6211120017	JD DE 900ML BOTTLES 25-45kGy	599	B/No.W031005 CUSTOMER B/No.3039
			ACTUAL DOSE35.4 - 37.0kGy
B2FLE0010999 B211120017	SAMPLES	5	
			CUSTOMER BATCH/CODE No.s AS ORDER No.P017811 DATED 23/11/2011
	Despatch Text:		
	12PLTS		
	Tofaï	604	Last Page 1 of 1

All in accordance with current Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



Wickham Laboratories Limited Winchester Road, Wickham, Fareham Hampshire PO17 5EU England Telephone: +44 (0)1329 832511

12 Dec 11

15 Dec 11

15 Dec 11

17885

Fax: +44 (0)1329 834262 mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Date Received:

Date Completed:

Purchase Order:

Date Tested:

CERTIFICATE OF ANALYSIS

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8

Hightown

White Cross Industrial Estate

Lancaster LA14XS

Test Article:

ClearKlens DE Block 1

Sample Code:

3039

Batch Ref: Qty Received:

WO31005 1 x 900mL Bottle

Laboratory Reference:

0002989/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Test	Method Item	Result	
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL	
Spiked Recovery	MM110/01	85 %	
Dilution Tested	MM110/04	1/50	
Endotoxin Spike Level	MM110/02	0.5 EU/mL	
Product Standard Data Sheet	MM110/05	FM01	

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

m Wood

Date: 16 Dec 11 14:13:03

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002989

Certificate of Analysis - OSMM

Print Number

P0000919

Lancaster

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



Wickham Laboratories Limited, Winchester Road, Wickham, Fareham, Hampshire PO17 5EU England Telephone: +44(0)1329 832511 +44(0)1329 834262 Fax:

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Date Received:

12 Dec 11

Flexible Medical Packaging Limited

Date Tested:

16 Dec 11

Unit 8

Mr J Jackson

Date Completed:

30 Dec 11

Hightown

White Cross Industrial Estate

Purchase Order:

17885

Lancaster LA14XS

Test Article:

ClearKlens DE Block 1

Sample Code:

3039 2013/11

Batch Ref:

WO31005

Qty Received:

20 x 900mL Bottles

Laboratory Reference:

0002989/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35° C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM03
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Even Wash

03 Jan 12 16:20:09 Date:

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

0002989

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment:

Certificate of Analysis - OSMM

Print Number

P0001078

Lancaster

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited