



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 13/1/12

Product Name	CLEARLENs TEGO 2000SC VH25
Product Code	7513387
Batch Number	W031197
Production Date	12/2011
Expiration Date	12/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		clear slightly yellow liquid
pH 100% (20°C)	JDM001	7.5	8.5	7.7
Specific Gravity (20°C)	JDM004	0.993	1.003	1.001

On behalf of Diversey site Quality Manager	Name :	C. Masoe
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 03	Date of issuing : November 28 <sup>th</sup> 2011
-------------	--------------	--

# CERTIFICATE OF IRRADIATION

Number: **B**

0515507 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster



## Synergy Health plc

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@synergyhealthplc.com  
Website: www.synergyhealthplc.com

LA1 4XS

P017943

Cust. Ref: P017943  
Date Rec'd 22/12/11  
Date: 3/01/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B211120804	DIVERSEY TEGO 2000SC 50ml 25-42kGy	104	B/No.W031197 - CUSTOMER B/No.387841  ROUTINE DOSE 34.6, 35.0 kGy
B2FLE0010999 B211120804	SAMPLES	1	  CUSTOMER BATCH/CODE No.5 AS ORDER No.P017943 DATED 21/12/11
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
Despatch Text:  2PLTS			
Total		105	Last Page 1 of 1
All in accordance with current			

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

Date Received: 16 Jan 12  
Date Tested: 18 Jan 12  
Date Completed: 06 Feb 12  
Purchase Order: 18007

Test Article: Clearklens Tego 2000 SC VII25  
Sample Code: WO31197  
Batch Ref: 387841  
Qty Received: 20 x 50mL and 5 spare bottles  
Laboratory Reference: 0003238/1  
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: P11 Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Approval is provided by Electronic Signature. Their name is shown below.

Date: 07 Feb 12 10:13:01

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003238  
Lancaster

Certificate of Analysis - OSMM

Print Number

P0001458

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*