Diversey

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20/1/12

Product Name	CLEARKLENS IPA VH1		
Product Code	7514372		
Batch Number	WO 31295		
Production Date	01/2012		
Expiration Date	01/2014		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower -	mits Upper	Results
Appearance	Visual		olourless quid	COLUMIESS
Specific Gravity (20°C)	JDM004	0.872	0.883	0.873

On behalf of Diversey site Quality Manager	Name :	Mise
	Position	Quality Control Inspector

CERTIFICATE OF IRRADIATION

Number: B Page

FLEU01

Flexible Medical Packaging Ltd

Unit 8

White Cross Industrial Estate

Hightown Lancaster our work protects your world

LA1 4XS

P018025

Cust. Ref: Date Rec'd 16/01/12

18/01/12

Date:

Synergy Health plc

Roydsdale Way · Euroway Trading Estate

Bradford · West Yorkshire BD4 6SE · United Kingdom Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061 Email:

assistance@synergyhealthplc.com Website: www.synergyhealthplc.com

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081	JD IPA BOTTLES 25-45kGy	689	B/No.W031295 - PRODUCT REF.7514372
B212010343			ROUTINE DOSE 39.3, 36.5, 37.5, 37.2 kGy
B2FLE0010999 B212010343	SAMPLES	6	
			CUSTOMER BATCH/CODE No.s AS ORDER No.P018025 DATED 16/1/12
	Despatch Text:		
	13PLTS		
	Total	695	Last Page 1 of 1
		All in acco	andance with aument

All in accordance with current Technical Agreement.

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories Limited. Winchester Road, Wickham, Fareham, Hampshire PO17 5EU England Telephone: +44(0)1329 832511

Fax:

+44(0)1329 834262

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS Date Received:

24 Jan 12

Date Tested:

27 Jan 12

Date Completed: Purchase Order:

27 Jan 12 18054

Test Article:

ClearKlens IPA Block 2

Sample Code:

3061

Batch Ref:

WO31295

Qty Received:

1 x 900mLs 0003323/2

Test Required:

Laboratory Reference:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Test	Method Item	Result	
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL	
Spiked Recovery	MM110/01	84 %	
Dilution Tested	MM110/04	1/40	
Endotoxin Spike Level	MM110/02	0.5 EU/mL	
Product Standard Data Sheet	MM110/05	FM02	

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Show Wash

30 Jan 12 16:40:13

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003323

Certificate of Analysis - OSMM

Print Number

P0001358

Lancaster

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



Wickham Laboratories Limited Winchester Road, Wickham, Fareham Hampshire PO17 5EU England Telephone: +44 (0)1329 832511

Fax: +44 (0)1329 834262 mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

24 Jan 2012

Date Tested:

26 Jan 2012

Date Test Completed:

09 Feb 2012

Purchase Order:

18054

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0003323/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

24/01/2012

Test Article:

ClearKlens IPA Block 2

Sample Code:

3061

Batch Ref:

WO31295

Qty Received:

20 x 900mLs + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM03
Volume Tested	MM107/05	20 X 50 mL
Comments		

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Eur Week

Date: 16 Feb 2012 18:36:17

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003323

Certificate of Analysis - OSMM

Print Number: P0001579

Page 1 of 1