

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20/2/12

Product Name	CLEARKLENS DE VH29		
Product Code	6069643		
Batch Number	WO 31433		
Production Date	02/2012		
Expiration Date	02/2014		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Colourless		CRIV adourto
	Lic		uid	1190x
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name:	Colose
	Position	Quality Control Inspector

This document being issued elec	stronically does not bear a signature	
COA 6069643	Version: 03	Date of issuing : November 28th 2011

CERTIFICATE OF IRRADIATION

Number: B 0520266 00

FLE001

Flexible Medical Packaging Ltd

Unit 8

White Cross Industrial Estate

Hightown Lancaster strial Estate
Synergyheath

our work protects your world

LA1 4XS

Cust. Ref:

P018143 9/02/12

Date Rec'd

13/02/12

Page 1

Synergy Health plc

Roydsdale Way · Euroway Trading Estate

Bradford · West Yorkshire BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011 Facsimile: +44 (0)1274 686061

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

Date: 13/02/	12		
ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B212020392	JD DE 900ML BOTTLES 25-45kGy	602	B/No.W031433 - CUSTOMER B/No.3258
			ROUTINE DOSE
B2FLE0010999 B212020392	SAMPLES	5	
			CUSTOMER BATCH/CODE No.s AS ORDER No.P018143 DATED 8/2/12
	Despatch Text:		
	12PLTS		
	Total	607	Last Page 1 of 1
		A11 :	

All in accordance with current Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories Limited Winchester Road, Wickham, Fareham Hampshire PO17 5EU England Telephone: +44 (0)1329 832511

20 Feb 2012

23 Feb 2012

23 Feb 2012

18195

Fax: +44 (0)1329 834262 mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Date Received:

Date Test Completed:

Date Tested:

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

Lancaster LA14XS

White Cross Industrial Estate

Purchase Order:

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0003645/3

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

20/02/2012

Test Article:

ClearKlens DE Block 1

Sample Code:

3258

Batch Ref:

Qty Received:

Works Order No. 31433

1 x 900mL

Method Item	Result
MM110/00	<0.2500 EU/mL
MM110/01	97 %
MM110/04	1/50
MM110/02	0.5 EU/mL
FG047/psd	FM01
	MM110/00 MM110/01 MM110/04 MM110/02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Show Wash

Date: 24 Feb 2012 13:34:32

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003645

Certificate of Analysis - OSMM

Print Number: P0001741

Page 1 of 1





Wickham Laboratories Limited, Hoeford Point, Barwell Lane, Gosport Hampshire PO13 OAU England

Telephone: +44(0)1329 226600

+44(0)1329 832511

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

20 Feb 2012

Date Tested:

23 Feb 2012

Date Test Completed:

08 Mar 2012

Purchase Order:

18195

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0003645/2

Test Required:

Date Received:

Sterillty Test by Membrane Filtration (Steritest) Method 1: Ph Eur and USP

20/02/2012

Test Article:

ClearKiens DE Block 1

Sample Code:

3258

Batch Ref:

Works Order No. 31433

Qty Received:

20 x 900mLs

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03
Commonte	1.5	

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 14 Mar 2012 08:35:37

Mrs P. Pham-Lengoc, BSc, MSB

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003645

Certificate of Analysis - OSMM

Print Number: P0001996

Page 1 of 1