



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 29/2/12

Product Name	CLEARLENSE CLEANSINALD SC VH9
Product Code	7516429
Batch Number	W031487
Production Date	03/2012
Expiration Date	03/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	12.0	13.0	12.7
Specific Gravity (20°C)	JDM004	1.040	1.060	1.041

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7516429	Version : 03	Date of issuing : November 28 <sup>th</sup> 2011
-------------	--------------	--

**CERTIFICATE OF IRRADIATION**Number: **B**

0521165 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster

**Synergy Health plc**

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

Cust. Ref: P018178  
Date Rec'd 16/02/12  
Date: 20/02/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010134 B212020608	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy	10	B/No.W031483 W031487 .
B2FLE0010999 B212020608	SAMPLES	1	
	Despatch Text:  ZPLTS		CUSTOMER BATCH/CODE No.s AS ORDER No.P018178 DATED 15/2/12
	Total	11	Last Page 1 of 1

All in accordance with current  
Technical Agreement

This is to certify that the above items have been irradiated as  
specified above

Authorising Signature: 

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 02 Mar 2012  
**Date Tested:** 07 Mar 2012  
**Date Test Completed:** 21 Mar 2012  
**Purchase Order:** 18261

### CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0003792/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP  
**Date Received:** 02/03/2012  
**Test Article:** Clearklens Cleansinald SC Block 1  
**Sample Code:** VLF 99C0110 340  
**Batch Ref:** WO31487  
**Qty Received:** 20 x 25mL bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 25 mL
Product Standard Data Sheet	FG047/psd	FM06

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 27 Mar 2012 14:45:03

Mrs P. Pham-Lengoc, BSc, MSB  
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003792

Certificate of Analysis - OSMM

Print Number: P0002190

Page 1 of 1