



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11/11/13

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP13288 34577
Production Date	10/2013
Expiration Date	08/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		clear slightly yellow liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.3
Specific Gravity (20°C)	JDM004	0.980	1.020	0.998

On behalf of Diversey site Quality Manager	Name :	Matthias O. Kirchner
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513403	Version : 04	Date of issuing : April 19 th 2012
-------------	--------------	---



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 02-Nov-2013

UK32S11066199-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002758
Customer Reference Number:	P021209
Product Description:	TEGO 2001 SS 25-45kGy
Validation Reference:	4.0968
Quantity Received:	133
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	LOT 4, 132 x FMP13288 34577 + 1 SAMPLE, 3 PLTS, THUR, 24.10.13

Irradiation Data

Date and Time of Irradiation:	02-Nov-2013 13:19
Reference Dose Range kGy:	35.3 - 35.8
Calculated Minimum Dose kGy:	25.2
Calculated Maximum Dose kGy:	38.6

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069