

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27/6/14

Product Name	CLEARKLENS CLEANSINALD SS VH9S	
Product Code	7516431	
Batch Number	FMP14178 36356	
Production Date	06/2014	
Expiration Date	12/2015	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower -	nits Upper	Results
Appearance	Visual		htly Yellow	Clear Stightly
pH (100% 20°C)	JDM001	9.0	11.4	11.3
Specific Gravity (20°C)	JDM004	0.990	1.010	0.993

On behalf of Diversey site Quality Manager	Name :	Miscoe	Mature
	Position	Quality Contr	

onically does not bear a signature	
Version: 04	Date of issuing: November 28th 2011
	onically does not bear a signature Version : 04



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 29-May-2014

UK32S11162723-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002720
Customer Reference Number:	P022264
Product Description:	C/SINALD SS 900ml BOTTLE 25-40kGy
Validation Reference:	4.0615
Quantity Received:	5
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	FMP B/NO: W036355 THURS 15.05.14 1PLT LOT 2

Irradiation Data

Date and Time of Irradiation:	27-May-2014 20:32	
Reference Dose Range kGy:	33.5 - 33.8	
Calculated Minimum Dose kGy:	27.6	
Calculated Maximum Dose kGy:	34.9	

Items irradiated under WO36355 will be used in finished product batch code

FMP14178 36356

K. Garaja . 30.06.14



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephone: +44(0)1329 832511

+44(0)1329 226600

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

01 Jul 2014

Date Tested:

03 Jul 2014

Date Test Completed:

17 Jul 2014

Purchase Order:

22483

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0017394/2

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

01/07/2014

Test Article:

Clearklens Cleansinald SS Block

Sample Code:

Batch Ref:

FMP14178 36356

Qty Received:

20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 21 Jul 2014 14:20:30

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0017394

Certificate of Analysis - OSMM

Print Number: P0021362

Page 1 of 1

