

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

## **CERTIFICATE OF ANALYSIS**

Date: 7/7/14

Product Name	CLEARKLENS IPA VH1	
Product Code	7514372	
Batch Number	FMP14167 36275	
Production Date	06/2014	
Expiration Date	06/2016	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	<b>Test Method</b>	Limits		Results
		Lower -		
Appearance	Visual	Clear C	olourless	Clear Colourless
•		Liquid		1,201d
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

On behalf of Diversey site  Quality Manager	Name:	CRESCOE AMaturo
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature			
COA 7514372	Version: 03	Date of issuing: November 28th 2011	



http://www.synergyhealthplc.com

# **Certificate of Irradiation**

Date Issued: 11-Aug-2014

UK32S11187332-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

#### Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1002734 **Customer Reference Number:** P022433 **Product Description:** JD IPA BOTTLES 25-45kGy Validation Reference: 4.0621 **Quantity Received:** 707 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 45.0 Customer Unit Lot/Batch Number: 703 X B/NO.FMP14167 36275 + 4 SAMPLES MON 23.06.14 13 PLTS **Irradiation Data** Date and Time of Irradiation: 02-Jul-2014 20:00 Reference Dose Range kGy: 35.3 - 40.0 Calculated Minimum Dose kGy: 25.5 Calculated Maximum Dose kGy: 42.5



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

07 Jul 2014

Date Tested:

11 Jul 2014

**Date Test Completed:** 

25 Jul 2014

**Purchase Order:** 

PO22537

#### **CERTIFICATE OF ANALYSIS**

Laboratory Reference Number:

0017503/3

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

**Date Received:** 

07/07/2014

Test Article:

Clearklens IPA Block 2

Sample Code:

-

Batch Ref: Qty Received: FMP14167 36275 21 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

#### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

amore

Date: 28 Jul 2014 13:51:50

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0017503

Certificate of Analysis - OSMM

Print Number: P0021558

Page 1 of 1

ISO





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

07 Jul 2014

Date Tested:

10 Jul 2014

**Date Test Completed:** 

10 Jul 2014

**Purchase Order:** 

22537

## **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0017503/4

**Test Required:** 

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

07/07/2014

Test Article:

Clearklens IPA Block 2

Sample Code:

5589

**Batch Ref:** 

FMP14167 36275

**Qty Received:** 

1 Bottle

Test	Method Item	Result	0)0484
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL	
Spiked Recovery	MM110/01	74 %	
Dilution Tested	MM110/04	1/40	
Endotoxin Spike Level	MM110/02	0.5 EU/mL	
Product Standard Data Sheet	FG047/psd	FM02	

#### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 11 Jul 2014 15:25:35

Mr K.A. Barker

Business Manager - Microbiology

Consignment: 0017503

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0021230

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire