



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 21.10.14

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	FMP 14267 36828
Production Date	09/2014
Expiration Date	09/2016

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

On behalf of Diversey site Quality Manager	Name :	A. Bouys	M. Matena
	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 09-Oct-2014

UK32S11252149-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	P023062
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1723
Quantity Received:	709
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	705 X B/N FMP14267 36828 + 4 SAMPLES MON 06.10.14 13 PLTS

Irradiation Data

Date and Time of Irradiation:	09-Oct-2014 19:35
Reference Dose Range kGy:	36.0 - 38.6
Calculated Minimum Dose kGy:	25.4
Calculated Maximum Dose kGy:	41.3

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Alt. Telephone: +44(0)1329 832511
Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 20 Oct 2014
Date Tested: 22 Oct 2014
Date Test Completed: 05 Nov 2014
Purchase Order: PO23135

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0019272/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 20/10/2014
Test Article: CLEARLENs IPA BLOCK 1
Sample Code: 5938
Batch Ref: FMP 14267 36828
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 06 Nov 2014 15:33:36

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0019272

Certificate of Analysis - OSMM

Print Number: P0023655

Page 1 of 1

ISO
17025

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited,
Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Alt. Telephone: +44(0)1329 832511
Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 20 Oct 2014
Date Tested: 21 Oct 2014
Date Test Completed: 21 Oct 2014
Purchase Order: PO23135

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0019272/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 20/10/2014
Test Article: CLEARLENs IPA BLOCK 1
Sample Code: 5938
Batch Ref: FMP 14267 36828
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	84 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 22 Oct 2014 15:27:45

Mr K.A. Barker
Business Manager - Microbiology

Consignment: 0019272

Certificate of Analysis - OSMM

Print Number: P0023303

Page 1 of 1

ISO
17025

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited,
Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire