

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 10/11/14

Product Name	CLEARKLENS IPA VH1	
Product Code	7513400	
Batch Number	FMP14298 37317	
Production Date	10/2014	
Expiration Date	10/2016	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	acronde.		Results
Α		Lower -	Upper	*
Appearance	Visual	Clear Colourless Liquid		Clear Colourles
				Lizurd
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875
				17:07
		1		

On behalf of Diversey site Quality Manager	Name:	Classe A.Bonys
	Position	Quality Control Inspector

Inis document being issued e	electronically does not bear a signature	
COA 7513400	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthpic.com

Certificate of Irradiation

Date Issued: 02-Nov-2014

UK32S11268598-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information		
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100432 1002734 P023201 JD IPA BOTTLES 25-45kGy 4.1723 178 25.0 45.0 174 X B/N FMP14298 37317 + 4 SAMPLES THUR 30.10.14 4 PLTS LOT 2	
II	rradiation Data	
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	01-Nov-2014 23:17 36.2 - 38.4 25.5 41.0	

Irradiation Release Authorised By Synergy Health plc



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England Telephone: +44(0)1329 226600

Alt. Telephone:

+44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

10 Nov 2014

Date Tested:

14 Nov 2014

Date Test Completed:

28 Nov 2014

Purchase Order:

23238

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0019655/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

10/11/2014

Date Received: Test Article:

Clearklens IPA Block 1

Sample Code:

_

Batch Ref:

FMP14298 37317

Qty Received:

20 x 900mL bottles 1 x spare

Method Item	Result
MM107/00	Pass
MM107/01	No growth in one broth
MM107/02	No growth in one broth
MM107/05	20 x 50 mL
FG047/psd	FM03
	MM107/00 MM107/01 MM107/02 MM107/05

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

lips

Date: 01 Dec 2014 15:21:52

Mrs P. Pham-Lengoc

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0019655

Certificate of Analysis - OSMM

Print Number: P0024118

Page 1 of

ISO 17025





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephone: +44(0)1329 226600 +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

10 Nov 2014

Date Tested:

12 Nov 2014

Date Test Completed:

12 Nov 2014

Purchase Order:

23238

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0019655/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

10/11/2014

Test Article:

Clearklens IPA Block 1

Sample Code:

Batch Ref:

Test

FMP 14298 37317 1 x 900mL bottle

Qty Received:

Method Item Result Bacterial Endotoxin by Kinetic QCL Test MM110/00 <0.2 EU/mL

spiked kecovery	MM110/01	96 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Ohn Wood

Date: 13 Nov 2014 16:29:27

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0019655

Certificate of Analysis - OSMM

Print Number: P0023784

Page 1 of 1

ISO 17025

