Diversey

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11/12/14

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP14318 37239
Production Date	11/2014
Expiration Date	09/2015 -

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

al CI	lear Sligh Ligi	ntly Yellow	gelow 1, quid
	LIG	uiu	1900
01 8	8.0	9.0	8.7
04 0.	.980	1.020	0.998
_			

On behalf of Diversey site Quality Manager	Name:	Mislar Mature
	Position	Quality Control Inspector

This document being issued electronica	lly does not bear a signature	•	
COA 7513403	Version: 04	Date of issuing: April 19th 2012	-



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 17-Dec-2014

UK31S11290020-1-1

This is to certify that Elgin Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Synergy Intersite - Bradford Roydsdale Way, Euroway Trading Estate Bradford Yorkshire BD4 6SE UNITED KINGDOM

	5 9	
Order	Informati	nn

Account Number:	UK32
Synergy Health Sales Part Reference:	1087099
Customer Reference Number:	S11281594
Product Description:	TEGO 2001 SS 2.0-7.0kGv
Quantity Received:	131
Customer Minimum Specification kGy:	2.0
Customer Maximum Specification kGy:	7.0
Customer Unit Lot/Batch Number:	SINGLE VALIDATION, DOSE REQ; 2.0-7.0 kGy, LOT FMP14318 37239, 340X280X220 @ 6.6KG
Other Process Details:	TOP UP FOR S11281594, VALIDATION REF NO
	5.1017 ACTUAL MIN:2.2 MAX: 3.0KGY
	Irradiation Data
Date and Time of Irradiation:	04-Dec-2014 05:21

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 24-Dec-2014

UK32S11281594-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002758
Customer Reference Number:	P0 23331
Product Description:	TEGO 2001 SS 25-45kGy
Validation Reference:	4.0968
Quantity Received:	131
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	129 X FMP14318 37239 + 2 SAMPLES THURS
	20.11.14 3PLTS LOT 3
Other Process Details:	Product originally received an under-dose of 1.0 kGy
	(as shown below) and was subsequently sent for
	dose augmentation at the Elgin Site 04 Dec 2014,
	NC ref 541. Actual augmented dose recieved =
	minimum 2.2 kGy and maximum 3.0 kGy
	(S11290020). Total certifiable dose = minimum
	26.2 kGy and maximum 41.0 kGy
lr.	rradiation Data

23-Nov-2014 05:12

Reference Dose Range kGy:

Calculated Minimum Dose kGy:

Calculated Maximum Dose kGy:

33.6 - 35.2

24.0

38.0

Irradiation Release Authorised By Synergy Health plc

Date and Time of Irradiation: