

Divorsey Europe BV

Maarasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 8427

CERTIFICATE OF ANALYSIS

Date: 27.04.2015

Product Name	CLEARKLENS CLEANSINALD SC VH9
Product Code	7516429
Batch Number	PMP 15117 32164
Production Date	3010412015
Expiration Date	30/104/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Slig	htly Yellow	clean stiphtly
		Liq	uid	Helias Liquid
pH (100% 20°C)	JDM001	12.0	13.0	12.2
Specific Gravity (20°C)	JDM004	1.040	1.060	1.050
	35/1004	1.040	1,000	1.050

On behalf of Diversey site Quality Manager	Name:	Opa Livenner, Mahue
	Position	Quality Control Inspector

This document being issued elect	tronically does not bear a signature	
COA 7516420	Various 102	170

COA 7516429 Version: 03 Date of issuing: November 28th 2011



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

06 May 2015

13 May 2015

27 May 2015

24287

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 **Hightown**

Lancaster LA14XS

White Cross Industrial Estate

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0022492/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

Purchase Order:

Date Test Completed:

Date Tested:

Date Received: 06/05/2015

Test Article:

ClearKlens Cleansinald SC

Sample Code:

Batch Ref: Qty Received: FMP15117 38164 20 x 25mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 25 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 28 May 2015 12:55:42

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0022492

Certificate of Analysis - OSMM

Print Number: P0027440

Page 1 of 1





http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 22-Apr-2015

UK32511372924-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order information			
Account Number:	100432		
Synergy Health Sales Part Reference:	1002719		
Customer Reference Number:	P024180		
Product Description:	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy		
Validation Reference:	4.1666		
Quantity Received:	11		
Customer Minimum Specification kGy:	25.0		
Customer Maximum Specification kGy:	40.0		
Customer Unit Lot/Batch Number:	11 X W038163 THURS 16 APR 2015 2 PLTS LOT 2		
I	rradiation Data		
Date and Time of Irradiation:	22-Apr-2015 15:56		
Reference Dose Range kGy:	33.5 - 34.0		
Calculated Minimum Dose kGy:	27.1		
Calculated Maximum Dose kGy:	34.9		

Items irradiated under WO38163 will be used in finished product batch code FMP15117 38164

