

Diversey Europe BV

Macrosenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tei. +31 (0)30 247 6427

# **CERTIFICATE OF ANALYSIS**

Date: 19.05.15

Product Name	CLEARKLENS IPA VH1		
Product Code	7514372		
Batch Number	PHP 15094 34911		
<b>Production Date</b>	16/04/2015		
<b>Expiration Date</b>	16 /04/2017		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Co	olourless	clean Calauntess
		Liq	uid	Liquiol
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875
	-			
	<del></del>		(8)	

On behalf of Diversey site  Quality Manager	Name:	Partycille, O. Kirchna
	Position	Quality Control Inspector

This document being issued electronical	lly does not bear a signature	
CO4 7514372	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

# **Certificate of Irradiation**

Date Issued: 11-May-2015

UK32S11374655-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

#### **Order Information Account Number:** 100432 Synergy Health Sales Part Reference: 1002734 **Customer Reference Number:** P024241 **Product Description:** JD IPA BOTTLES 25-45kGy Validation Reference: 4.1723 710 **Quantity Received: Customer Minimum Specification kGy:** 25.0 **Customer Maximum Specification kGy:** 45.0 Customer Unit Lot/Batch Number: B/NO: 706 X FMP15097 37911 + 4 SAMPLES -THURS 23 APR 2015 13PLTS LOT 1 Other Process Details: Released against validation 4.1971. Calculated minimum dose 26.5 kGy and calculated maximum dose 40.7 kGy. Irradiation Data Date and Time of Irradiation: 08-May-2015 15:17



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

13 May 2015

Date Tested:

20 May 2015

Date Test Completed:

20 May 2015

Purchase Order:

24343

### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0022637/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

13/05/2015

Test Article:

**IPA** 

Sample Code: Batch Ref:

FMP15097 37911

Qty Received:

1 x 900ml Bottles

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	87 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

#### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Sun Daw

Date: 21 May 2015 17:47:48

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0022637

Certificate of Analysis - OSMM

Print Number: P0027354

Page 1 of 1





Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alf. Telephone: +44(0)1329 832511 Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

13 May 2015

Date Tested:

20 May 2015

Date Test Completed:

03 Jun 2015

Purchase Order:

24343

#### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0022637/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received: Test Article: 13/05/2015 ClearKlens IPA

Sample Code:

CIG

Batch Ref:

FMP15097 37911

Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

## **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Eun Ward

Date: 04 Jun 2015 17:19:10

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0022637

Certificate of Analysis - OSMM

Print Number: P0027587

Page 1 of 1

ISQ 17025

