

Diversey Europe BV

Maarasenbroeksedijk 2 3542 DN Utrocht The Netherlands

Tel. +31 (0)30 247 6427

## **CERTIFICATE OF ANALYSIS**

Date: 4.6.15

Product Name	CLEARKLENS IPA VH1	
Product Code	7513400	
Batch Number	FNP15117 38162	
Production Date	27/04/2015	
Expiration Date	27/04/2017	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	<b>Test Method</b>	Li	mits	Results
		Lower -	Upper	
Appearance	Visual	Clear C	olourless	clear Colourless
		Lie	guid	Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875
				<del></del>

On behalf of Diversey site Quality Manager	Name:	Parlying Mature
	Position	Quality Control Inspector

1 his accument being issued electronical	y does not bear a signature	
COA 7513400	Yersion: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

# **Certificate of Irradiation**

Date Issued: 22-May-2015 UK32S11386153-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

#### **Order Information** Account Number: 100432 **Synergy Health Sales Part Reference:** 1002734 **Customer Reference Number:** P024290 JD IPA BOTTLES 25-45kGy **Product Description:** Validation Reference: 4.1971 **Quantity Received:** 708 25.0 **Customer Minimum Specification kGy: Customer Maximum Specification kGy:** 45.0 704 X B/N FMP15117 38162 + 4 SAMPLES THUR **Customer Unit Lot/Batch Number:** 07 MAY 2015 13 PLTS LOT 2 Irradiation Data Date and Time of Irradiation: 22-May-2015 19:40 37.5 - 38.3 Reference Dose Range kGy: 27.0 Calculated Minimum Dose kGy: Calculated Maximum Dose kGy: 39.8



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England Telephone: +44(0)1329 226600

Telephone: Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

03 Jun 2015

Date Tested:

10 Jun 2015

Date Test Completed:

24 Jun 2015

Purchase Order:

24454

### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0022954/2

Test Required:
Date Received:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

03/06/2015

Test Article:

ClearKlens IPA Block 1

Sample Code:

6449

Batch Ref:

FMP 15117 38162

**Qty Received:** 

20 x 900mL Bottles + 1 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

#### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

UNERKOVE.

Date: 25 Jun 2015 10:05:44

Mrs C Moore

Page 1 of 1

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment; 0022954 Print Number: P0027943









Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

03 Jun 2015

Date Tested:

04 Jun 2015

Date Test Completed:

04 Jun 2015

Purchase Order:

24454

### **CERTIFICATE OF ANALYSIS**

Laboratory Reference Number:

0022954/4

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received: Test Article: 03/06/2015 IPA Block 1

6449

Sample Code: Batch Ref: Qty Received:

FMP 15117 38162

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	72 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

#### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Sun Uses

Date: 05 Jun 2015 16:03:25

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0022954

Certificate of Analysis - OSMM

Print Number: P0027604

Page 1 of 1

ISO 17025

