

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tei. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27.06.15

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	PHP 15162 38276
Production Date	11/06/2015
Expiration Date	11/04/12016

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Liu	mits	Results
		Lower -	Upper	
Appearance	Visual	Clear Slig	htly Yellow	clean slightly
		Lic		Yellow Lipuid
pH 100% (20°C)	JDM001	8.0	9.0	8.6
Specific Gravity (20°C)	JDM004	0.980	1.020	0.998

O-1-1-15-6Di	Name :	apadirchner, Malus
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513403 Version: 04 Date of issuing:	3 Version: 04	Date of issuing: April 19th 2012
------------------------------------------	---------------	----------------------------------



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 25-Jun-2015 UK32S11409833-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information				
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy:	100432 1002758 P024548 TEG0 2001 SS 25-45kGy 4.1903 132 25.0 45.0			
Customer Unit Lot/Batch Number:	131 X B/N FMP15162 38276 + 1 SAMPLE THUR 18 JUN 2015 3 PLTS LOT 2			
I	rradiation Data			
Date and Time of Irradiation:	24-Jun-2015 21:47			
Reference Dose Range kGy:	35,4 - 38.4			
Calculated Minimum Dose kGy:	25.0			
Calculated Maximum Dose kGy:	40.2			