

ACTIVATOR
FIRST OFF



Diversey Europe Operations BV

Maarssebroekse dijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 14.04.2015

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7518162
Batch Number	FMP 15104 38308
Production Date	14/04 / 2015
Expiration Date	14/04 / 2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (100% 20°C)	JDM001	9.0	12.5	11.5
Specific Gravity (20°C)	JDM004	1.004	1.020	1.006
Potential Chlorine Dioxide (ppm)		100	180	141.75ppm

On behalf of Diversey site Quality Manager	Name :	Alpa Kishna, Parthasarathy
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Activator	Version : 01	Date of issuing : October 1 st 2012
------------------------	--------------	--

BASE
FIRST OFF



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 14.04.2015

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7518162
Batch Number	FMP 15104 38308
Production Date	14/04/2015
Expiration Date	14/04/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (100% 20°C)	JDM001	1.5	2.5	1.7
Specific Gravity (20°C)	JDM004	1.010	1.030	1.018

On behalf of Diversey site Quality Manager	Name :	Olga Kinnear, Parvisha
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Base	Version : 01	Date of issuing : September 27 th 2012
-------------------	--------------	---



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 20-May-2015

UK32S11387809-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process
In accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002735
Customer Reference Number:	PO24360
Product Description:	JD-BOTTLES/CAPS/BAGS 25-40kGy
Validation Reference:	4.1624
Quantity Received:	8
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	B/N WO38307 THUR 14 MAY 2015 2 PLTS LOT 3

Irradiation Data

Date and Time of Irradiation:	20-May-2015 10:15
Reference Dose Range kGy:	34.7 - 35.0
Calculated Minimum Dose kGy:	28.8
Calculated Maximum Dose kGy:	36.6

Items irradiated under WO38307 will be used in finished product batch code FMP15104 38308

E. Gargan 10.06.15

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 11 Jun 2015
Date Tested: 17 Jun 2015
Date Test Completed: 01 Jul 2015
Purchase Order: PO24503

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0023091/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 11/06/2015
Test Article: Bi-Spore Activator
Sample Code: 502304
Batch Ref: FMP15104 38308
Qty Received: 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM04

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 02 Jul 2015 16:20:00

Mr K.A. Barker
Business Manager - Microbiology



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 11 Jun 2015
Date Tested: 17 Jun 2015
Date Test Completed: 01 Jul 2015
Purchase Order: PO24503

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0023091/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 11/06/2015
Test Article: Bi-Spore Base
Sample Code: 502303
Batch Ref: FMP15104 38308
Qty Received: 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM05

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 02 Jul 2015 16:23:36

Mr K.A. Barker
Business Manager - Microbiology