



Diversey Europe BV
 Maarssebroeksedijk 2
 3542 DN Utrecht
 The Netherlands
 Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 7.09.2015

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	FHP 15171 38351
Production Date	20/06/2015
Expiration Date	20/06/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.879

On behalf of Diversey site Quality Manager	Name :	opa kinenna, Purbijiso
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

CCA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 24-Jul-2015

UK32S11418575-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	P024608
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1971
Quantity Received:	709
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	704 x B/N FMP15171 38351 + 5 SAMPLES TUES 30TH JUNE 2015 13 PLTS LOT 1

Irradiation Data

Date and Time of Irradiation:	24-Jul-2015 14:31
Reference Dose Range kGy:	37.2 - 39.3
Calculated Minimum Dose kGy:	26.8
Calculated Maximum Dose kGy:	40.8

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 10 Aug 2015
Date Tested: 19 Aug 2015
Date Test Completed: 02 Sep 2015
Purchase Order: PO24894

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024114/1
Test Required: Sterility Test by Membrane Filtration (Steritest)
Date Received: 10/08/2015
Test Article: ClearKlens IPA
Sample Code: -
Batch Ref: FMP 15171 38351
Qty Received: 20 Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 03 Sep 2015 15:55:46

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM
Consignment: 0024114
Print Number: P0029432
Page 1 of 1



Company Registered in England No. 752951



Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 10 Aug 2015
Date Tested: 12 Aug 2015
Date Test Completed: 12 Aug 2015
Purchase Order: PO24894

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024114/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 10/08/2015
Test Article: IPA
Sample Code: 6690
Batch Ref: FMP 15171 38351
Qty Received: 1 Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	88 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 13 Aug 2015 14:27:30

Mr K.A. Barker
Business Manager - Microbiology