

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 7 9 15

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP 15231 38638
Production Date	19/08/2015
Expiration Date	19/06/2016

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin	nits	Results
		Lower -	Upper	
Appearance	Visual	Clear Slig	htly Yellow	deat slightly,
		_	_l uid	Hellow Liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.8
Specific Gravity (20°C)	JDM004	0.980	1.020	0.998

On behalf of Diversey site Quality Manager	Name:	Butyingles, Meta	e
	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA 7513403	Version: 04	Date of issuing: April 19th 2012
-------------	-------------	----------------------------------



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 04-Sep-2015

UK32S11457571-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information			
Account Number:	100432		
Synergy Health Sales Part Reference:	1002758		
Customer Reference Number:	P024999		
Product Description:	TEGO 2001 SS 25-45kGy		
Validation Reference:	4.1903		
Quantity Received:	122		
Customer Minimum Specification kGy:	25.0		
Customer Maximum Specification kGy:	45.0		
Customer Unit Lot/Batch Number:	121 X B/N FMP15231 38638 THURS 27TH AUG		
	2015 2 PLTS LOT 2 + 1 SAMPLE		
lı lı	rradiation Data		
Date and Time of Irradiation:	04-Sep-2015 04:23		
Reference Dose Range kGy:	37.8 - 38.3		
Calculated Minimum Dose kGy:	26.7		
Calculated Maximum Dose kGy:	40.1		

Irradiation Release Authorised By Synergy Health plc