



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 7.9.15

Product Name	CLEARLENs IPA VH1
Product Code	7513400
Batch Number	FMP1520R 38637
Production Date	27/07/2015
Expiration Date	27/07/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

On behalf of Diversey site Quality Manager	Name :	Parry, Mafun
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 30-Aug-2015

UK32S11447593-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	PO24913
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1971
Quantity Received:	1420
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	1415 X B/N FMP15208 38637 + 5 SAMPLES THUR 13 AUG 2015 27 PLTS LOT 1

Irradiation Data

Date and Time of Irradiation:	30-Aug-2015 00:40
Reference Dose Range kGy:	36.9 - 39.0
Calculated Minimum Dose kGy:	26.6
Calculated Maximum Dose kGy:	40.5

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 07 Sep 2015
Date Tested: 09 Sep 2015
Date Test Completed: 23 Sep 2015
Purchase Order: 25064

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024569/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 07/09/2015
Test Article: Clearklens IPA
Sample Code: -
Batch Ref: FMP 15208 38637
Qty Received: 20 Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 24 Sep 2015 14:40:31

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 07 Sep 2015
Date Tested: 08 Sep 2015
Date Test Completed: 08 Sep 2015
Purchase Order: 25064

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024569/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 07/09/2015
Test Article: IPA
Sample Code: *
Batch Ref: FMP 15208 38637
Qty Received: 1 Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	103 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 08 Sep 2015 16:12:14

Mr K.A. Barker
Business Manager - Microbiology