



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 15.09.15

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	100848254
Batch Number	FMP15258 38948
Production Date	15/09/2015
Expiration Date	15/03/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear Slightly Yellow Liquid
pH (100% 20°C)	JDM001	9.0	11.4	10.8
Specific Gravity (20°C)	JDM004	0.990	1.010	0.994

On behalf of Diversey site Quality Manager	Name :	J. Staver	Matina
	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA 7516432	Version : 05	Date of issuing : April 27 th 2012
-------------	--------------	---



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 15-Oct-2015

UK32S11451300-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002720
Customer Reference Number:	P024951
Product Description:	C/SINALD SS 900ml BOTTLE 25-40kGy
Validation Reference:	4.1909
Quantity Received:	5
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	B/N WO38947 THURS 20TH AUG 2015 1 PLT LOT 6

Irradiation Data

Date and Time of Irradiation:	02-Sep-2015 12:00
Reference Dose Range kGy:	34.4 - 34.6
Calculated Minimum Dose kGy:	30.1
Calculated Maximum Dose kGy:	37.1

Items irradiated under WO38947 will be used in finished product number: FMP15258 38948

L. Gwyn 15.10.15

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 17 Sep 2015
Date Tested: 29 Sep 2015
Date Test Completed: 13 Oct 2015
Purchase Order: 25124

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024752/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 17/09/2015
Test Article: Clearklens Cleansinald ss
Sample Code: FMP 15258 38948
Batch Ref: 6893
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

Tested in accordance with Ph Eur 8.0 2.6.1, USP 38 <71> & JP XVI 4.06 except for volume tested.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 14 Oct 2015 15:41:49

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology