



Diversey Europe BV

Messegembroekdijk 2  
3842 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 20.11.2015

Product Name	CLEARLENs IPA VH1
Product Code	7513400
Batch Number	FMP15272 / 38951
Production Date	29.09.2015
Expiration Date	29.09.2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.879

On behalf of Diversey site Quality Manager	Name :	P. J. Staan
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 <sup>th</sup> 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 11-Nov-2015

UK32S11486996-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products  
EN ISO 9001:2008 Quality Management System  
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd  
Unit 8, White Cross Ind Estate  
Hightown  
Lancaster  
Lancashire LA1 4XS  
UNITED KINGDOM

All in accordance with current Technical Agreement

## Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	P025303
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1971
Quantity Received:	717
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	713 X FMP15272 38951 THURS 15TH OCT 2015 14 PLTS LOT 2

## Irradiation Data

Date and Time of Irradiation:	02-Nov-2015 01:38
Reference Dose Range kGy:	35.0 - 38.6
Calculated Minimum Dose kGy:	25.2
Calculated Maximum Dose kGy:	40.1

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 23-Oct-2015

UK32S11482847-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products  
EN ISO 9001:2008 Quality Management System  
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd  
Unit 8, White Cross Ind Estate  
Hightown  
Lancaster  
Lancashire LA1 4XS  
UNITED KINGDOM

All in accordance with current Technical Agreement

---

## Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	PO25250
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1971
Quantity Received:	702
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	BATCH NO: FMP15272 38951 THURS 8TH OCT 2015 13 PLTS LOT 3

---

## Irradiation Data

Date and Time of Irradiation:	23-Oct-2015 15:10
Reference Dose Range kGy:	36.9 - 39.9
Calculated Minimum Dose kGy:	26.6
Calculated Maximum Dose kGy:	41.4

---

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



# Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd  
Hoeford Point, Barwell Lane, Gosport  
Hampshire PO13 0AU England  
Telephone: +44(0)1329 226600  
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk  
www.wickhamlabs.co.uk

Ms D Henderson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 17 Nov 2015  
**Date Tested:** 20 Nov 2015  
**Date Test Completed:** 04 Dec 2015  
**Purchase Order:** PO25490

## CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0025867/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP  
**Date Received:** 17/11/2015  
**Test Article:** IPA  
**Sample Code:** FMP 15272 38951  
**Batch Ref:** 6995  
**Qty Received:** 20 x 900mL Bottles (+1 spare)

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

### Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 07 Dec 2015 12:09:14

Mrs C Moore  
Laboratory Manager - Pharmaceutical Microbiology



# Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd  
Hoeford Point, Barwell Lane, Gosport  
Hampshire PO13 0AU England  
Telephone: +44(0)1329 226600  
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk  
www.wickhamlabs.co.uk

Ms D Henderson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 17 Nov 2015  
**Date Tested:** 18 Nov 2015  
**Date Test Completed:** 18 Nov 2015  
**Purchase Order:** 25490

## CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0025867/2  
**Test Required:** Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP  
**Date Received:** 17/11/2015  
**Test Article:** Clearklens IPA  
**Sample Code:** FMP15272 38951  
**Batch Ref:** 6995  
**Qty Received:** 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	95 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 19 Nov 2015 11:18:05

Mrs P. Pham-Lengoc  
Laboratory Manager - Pharmaceutical Microbiology