



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20.11.2015

Product Name	CLEARKLENS TEGO 2000SS VH25S
Product Code	7516427
Batch Number	FMP 15301 39267
Production Date	28.10.2015
Expiration Date	28.04.2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
pH 100% (20°C)	JDM001	6.2	8.2	7.5
Specific Gravity (20°C)	JDM004	0.990	1.010	0.998

On behalf of Diversey site Quality Manager	Name :	M. Glau
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA VH25S	Version : 05	Date of issuing : June 11 th 2013
-----------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 13-Nov-2015

UK32S11501272-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1035147
Customer Reference Number:	P025436
Product Description:	TEGO 2000SS BLK1 25-45KGY
Validation Reference:	4.1903
Quantity Received:	62
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	BATCH NO. FMP15301 39267 THURS 5TH NOV 2015 1 PLT LOT 1 + 4 SAMPLES

Irradiation Data

Date and Time of Irradiation:	13-Nov-2015 17:46
Reference Dose Range kGy:	36.1 - 36.5
Calculated Minimum Dose kGy:	25.5
Calculated Maximum Dose kGy:	38.2

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 18 Nov 2015
Date Tested: 25 Nov 2015
Date Test Completed: 09 Dec 2015
Purchase Order: 25497

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0025896/3
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 18/11/2015
Test Article: Tego2000 ss
Sample Code: FMP 15301
Batch Ref: 39267
Qty Received: 21 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 10 Dec 2015 13:32:59

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology