



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 18.12.2015

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP 15821, 39269
Production Date	17/11/2015
Expiration Date	EXP 17/09/2016

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear Slightly Yellow Liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.4
Specific Gravity (20°C)	JDM004	0.980	1.020	0.998

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

*This document being issued electronically does not bear a signature*

COA 7513403	Version : 04	Date of issuing : April 19 <sup>th</sup> 2012
-------------	--------------	---



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 12-Dec-2015

UK32S11516569-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd  
Unit 8, White Cross Ind Estate  
Hightown  
Lancaster  
Lancashire LA1 4XS  
UNITED KINGDOM

All in accordance with current Technical Agreement

---

## Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002758
Customer Reference Number:	P025556
Product Description:	TEGO 2001 SS 25-45kGy
Validation Reference:	4.1903
Quantity Received:	134
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	133 X BATCH NO: FMP15321 39269 THURS 26TH NOV 2015 3 PLTS LOT 3

---

## Irradiation Data

Date and Time of Irradiation:	12-Dec-2015 02:30
Reference Dose Range kGy:	35.5 - 37.2
Calculated Minimum Dose kGy:	25.1
Calculated Maximum Dose kGy:	38.9

---

Irradiation Release Authorised By Synergy Health plc

---

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069