

Diversey Europe Operations BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 18.05.2017

CLEARKLENS CLEANSINALD SS VH9S
100848254
FHP 17 138 42468
18105/2017
Exp. 18/11/2018

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	٠.
Appearance	Visual	Clear Slig		clean sciently
		Lio	uid	tellow Liques
pH (100% 20°C)	JDM001	9.0	11.4	11.0
Specific Gravity (20°C)	JDM004	0.990	1.010	0.996
			2,020	0, 5196

On behalf of Diversey site Quality Manager	Name :	O. Linconne, Butgerst
	Position	Quality Control Inspector

This document being issued e	ectronically does not bear a signature	
COA 100848254	Version: 02	
	7 6/310/1 . 02	Date of issuing: April 7th 2016



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Fax:

+44(0)1329 226600 +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

25 May 2017

Date Tested:

02 Jun 2017

Date Test Completed:

16 Jun 2017

Purchase Order:

28994

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0035777/1

Test Required:

Sterility by Membrane Filtration Steritest

Date Received:

25/05/2017

Test Article:

ClearKlens Cleansinald SS

Sample Code:

FMP17138 42768

Batch Ref:

8469

Qty Received:

20 x 900mL Bottles + 1 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

BSC CBiol MRSB

Date: 16 Jun 2017 17:13:17

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0035777 Print Number: P0043543

Page 1 of 1





http://www.synergyhealthpic.com

Certificate of Irradiation

Date Issued: 08-May-2017

UK33S11867319-1-4

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

Calculated Maximum Dose kGy:

All in accordance with current Technical Agreement

Order Information Account Number: 100432 Synergy Health Sales Part Reference: 1105380 **Customer Reference Number:** P028853 **Product Description:** C/SINALD SS 900ml Bottle DV4673 25-40kGy Validation Reference: 4673 **Quantity Received: Customer Minimum Specification kGy:** 25.0 **Customer Maximum Specification kGy:** 40.0 W042769, 1 PLT Customer Unit Lot/Batch Number: **Irradiation Data** 08-May-2017 03:16 Date and Time of Irradiation: Reference Dose Range kGy: 31.9 - 32.0 Calculated Minimum Dose kGy: 29.3

35.0

Items irradiated under WO42769, will be used in finished product batch FMP17138 42768

Irradiation Release Authorised By Synergy Health plc