



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 12.07.2017

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP 17163 42856
Production Date	12/06/2017
Expiration Date	EXP 12/04/2018

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly Yellow Liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.4
Specific Gravity (20°C)	JDM004	0.980	1.020	1.000

On behalf of Diversey site Quality Manager	Name :	Pantjisto, V. M. C.
	Position	Quality Control Inspector

*This document being issued electronically does not bear a signature*

COA 7513403	Version : 04	Date of issuing : April 19 <sup>th</sup> 2012
-------------	--------------	---

<http://www.synergyhealthplc.com>

## Certificate of Irradiation

Date Issued: 24-Jun-2017

UK33S11902568-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products  
EN ISO 9001:2008 Quality Management System  
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd  
Unit 8, White Cross Ind Estate  
Hightown  
Lancaster  
Lancashire LA1 4XS  
UNITED KINGDOM

All in accordance with current Technical Agreement

### Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1106485
Customer Reference Number:	P029110
Product Description:	TEGO 2001SS DV4767 25-45kGy
Validation Reference:	4767
Quantity Received:	131
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	FMP17163 42856, INC 1 SAMPLE BOX, 3 PLTS

### Irradiation Data

Date and Time of Irradiation:	24-Jun-2017 22:44
Reference Dose Range kGy:	34.8 - 35.2
Calculated Minimum Dose kGy:	27.6
Calculated Maximum Dose kGy:	39.9

Irradiation Release Authorised By Synergy Health plc

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069