



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 15.12.2017

Product Name	CLEARLENSS TEGO 2000SS VH25S
Product Code	7516427
Batch Number	FMP 17326 44024
Production Date	22/11/2017
Expiration Date	EXP 22/05/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear Slightly Yellow Liquid
pH (neat solution)	DM001	6.2	8.2	7.6
Specific Gravity (20°C)	DM004	0.990	1.010	0.996

On behalf of Diversey site Quality Manager	Name :	Parkyeist, Navaburob
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 th 2017
--------------	--------------	--



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 03-Dec-2017

UK33S12021752-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1111557
Customer Reference Number:	P030042
Product Description:	TEGO 2000SS BLK1 DV4767 25-45KGY
Validation Reference:	4767
Quantity Received:	61
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	FMP17326 44024, INC 4 SAMPLES, 1 plt

Irradiation Data

Date and Time of Irradiation:	03-Dec-2017 12:17
Reference Dose Range kGy:	37.8 - 38.1
Calculated Minimum Dose kGy:	30.0
Calculated Maximum Dose kGy:	43.2

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 19 Dec 2017
Date Tested: 10 Jan 2018
Date Test Completed: 24 Jan 2018
Purchase Order: 30178

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0039908/1
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 19/12/2017
Test Article: Clearklens Tego 2000SS
Sample Code: FMP17326 44024
Batch Ref: 8985
Qty Received: 20 x 900mL Bottles + 2 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 40 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM08

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

CBIOL MKSB

Date: 25 Jan 2018 13:04:17

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology