Finst off



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 12.11.18

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	FMP 18312 46025
Production Date	08/11/2018
Expiration Date	Exp. 08/11/2020

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin	nits	Results
		Lower	- Upper	
Appearance	Visual	Clear Colou	ırless Liquid	acon Colorines
pH (neat solution)	DM001	9.0	12.5	4.4
Specific Gravity (20°C)	DM004	1.004	1.020	1.006
Potential Chlorine Dioxide (ppm)		100	180	145.125 ppm

	Name:	aga Linconne, Pourlywood
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature			
COA Template	Version: 02	Date of issuing: November 24th 2017	





Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 08/11/2018

Product Code 7514852 Batch Number FMP 18312 46025 Production Date 08/11/2018 Expiration Date Exp. 08/11/2020	Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Production Date 08/11/2018	Product Code	7514852
	Batch Number	FMP 18312 46025
Expiration Date Exp. 08 111 2020	Production Date	08/11/2018
	Expiration Date	Exp. 08/11/2020

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin	nits	Results
		Lower	- Upper	
Appearance	Visual		olourless	Clean Colourless
		Lic	quid	liquial
pH (neat solution)	DM001	1.5	2.5	1.6
Specific Gravity (20°C)	DM004	1.010	1.030	1.016

O-1-1-16-6D''	Name :	aga kirchner, Partyust
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version: 02	Date of issuing : November 24 th 2017
--------------	-------------	--



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

21 Nov 2018

Date Tested:

27 Nov 2018

Date Test Completed:

11 Dec 2018

Purchase Order:

32166

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0045940/2

Test Required:

Sterility by Membrane Filtration Steritest

Date Received: **Test Article:**

21/11/2018

Bi-Spore Activator Sample Code: FMP18312 46025

Batch Ref: Qty Received:

616965 20 x 100mL

	(TOOTTIE	
Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 41 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM04-01

Approval is provided by Bectronic Signature. Their name and position is shown below.

CBiol MRSB

Date: 13 Dec 2018 15:31:50

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0045940 Print Number: P0055741

Page 1 of 1





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600

Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 **Hightown**

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

21 Nov 2018

Date Tested:

27 Nov 2018

Date Test Completed:

11 Dec 2018

Purchase Order:

32166

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0045940/1

Test Required:

Sterility by Membrane Filtration Steritest

Date Received: Test Article:

21/11/2018 Bi-Spore Base

Sample Code: Batch Ref:

FMP18312 46025 622518

Qty Received: 20 x 100mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 ml
Tested in Accordance with Ph Eur, USP		
& JP	MQ005/07	EP 9.0 2.6.1, USP 41 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM05-01

Approval is provided by Electronic Signature. Their name and position is shown below.

SAPPLY CBiol MRSB

Date: <u>13 Dec 2018 15:31:50</u>

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0045940 Print Number: P0055741

Page 1 of 1





http://www.steris-ast.com

Certificate of Irradiation

Date issued: 27-Oct-2018

UK33S12184414-2-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 **Synergy Health Sales Part Reference:** 1106825 **Customer Reference Number:** P031918 **Product Description:** BISP B3 COMPONENTS DV4725 25-40kGv Validation Reference: 4725 **Quantity Received: Customer Minimum Specification kGy:** 25.0 Customer Maximum Specification kGy: 40.0 **Customer Unit Lot/Batch Number:** W046026, 1 PLT Irradiation Data Date and Time of Irradiation: 27-Oct-2018 00:10 Reference Dose Range kGy: 33.6 - 33.8 Calculated Minimum Dose kGy: 27.8 Calculated Maximum Dose kGy: 34.5

Items irradiated under WO46026 will be used in finished product batch FMP18312 46025

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company