

7441658

<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 14-Jan-2019

FR01S12218032-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

ECP SAS
Le Millénaire
395 rue Louis Lépine
34000 MONTPELLIER
FRANCE

DIVERSEY
Clearklem Bi-Spore RTU VH26S
Batch n° ENT 41229 19 080
Y. FERRAND
06/05/19

Order Information

Account Number:	101999
Synergy Health Sales Part Reference:	1126321
Customer Reference Number:	CF181551 - 21/12/2018
Product Description:	224521P - CONSOMMABLES BIDONS ET BOUCHONS <= 5L
Validation Reference:	S12126430
Quantity Received:	18
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Other Process Details:	The treatment of this batch was applied in two fractions.

Irradiation Data

Date and Time of Irradiation:	11-JAN-2019 16:45
Calculated Minimum Dose kGy:	25.0
Calculated Maximum Dose kGy:	38.6


Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1



7482164

<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Jan-2019

FR01S12221815-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

ECP SAS
Le Millénaire
395 rue Louis Lépine
34000 MONTPELLIER
FRANCE

DIVERSEY
Clearkless Bi-Spare RTU VH2GS
Batch n° ENT 41229 19 080
M. FERRAND
06/05/19

Order Information

Account Number:	101999
Synergy Health Sales Part Reference:	1126736
Customer Reference Number:	CF190039 - 08/01/2019
Product Description:	224522P - CONSOMMABLES Sacs, Etiquettes, Gants
Validation Reference:	S12131503
Quantity Received:	1
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	55.0
Other Process Details:	The treatment of this batch was applied in two fractions.

Irradiation Data

Date and Time of Irradiation:	26-JAN-2019 11:22
Calculated Minimum Dose kGy:	26.8
Calculated Maximum Dose kGy:	51.2

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1

Customer DIVERSEY LTD	Management Quality Manual n° MMQ0000 (contractual)
--------------------------	--

Product : Clearklens Bi-Spore RTU VH26S	Control report n° : CERT 19194
Specification reference : NA	Purchase order n° : 4702350591
Analyst : MJA	Date of receipt : 03/28/2019

PHYSICAL AND CHEMICAL PROPERTIES CONTROL

Customer's batch n° : ENT41229 19 080	Certificate Nb STERIS : S122180320101 S122218150101
Date of manufacturing : 03/2019	Produced quantity : 160 x 4,9L / 160 x 0,1L
Expiry date : 03/2021	FIDT n° : 1816 ind AB
ECP's batch n° : 41229	Date of analysis : 03/28/2019

Results :

Solution	Analyzed characteristics	Method of analysis	State in the production	Specification
			Beginning	
Bi-Spore Activator Solution	Appearance (20°C)	NA	Clear colourless liquid	Clear colourless liquid
	pH (20°C)	IDT 0312	11,9	9,0 - 12,5
	Specific gravity (20°C)	IDT 0301	1,008	1,004 - 1,020
Bi-Spore Base Solution RTU	Appearance (20°C)	NA	Clear colourless liquid	Clear colourless liquid
	pH (20°C)	IDT 0312	2,8	2,5 - 5,0
	Specific gravity (20°C)	IDT 0301	1,000	1,000 - 1,010
Bi-Spore Activator Solution + Base Solution RTU	Appearance (20°C)	NA	Clear yellow liquid	Clear yellow liquid
	pH (20°C)	IDT 0312	3,0	2,5 - 5,0
	Specific gravity (20°C)	IDT 0301	1,000	1,000 - 1,010
	Dosage of the material chlorine dioxide	IDT 0261	108	100 - 180 ppm
Microbiological contagion of the WFI		IDT 0236	9 cfu/100mL	< 10 cfu/100mL

C : Conform NC : Not Conform

NA : Not Applicable

Batch released by H. FERRAND 06/05/19

(Signature: H. Ferrand)

Conclusion :

☒

CONFORM

☐

NOT CONFORM

Date : 04/03/2019

Written by : M. JARRIN

Checked by :

M. FERRAND

IMP 0180 - L



MedicalGroup

11/10/2018, Approuvé par

Report N°:

19/OI/STE/034

Demandeur :

ECP

Enquêter :

395, rue Louis Léprie
34000 MONTPELLIER

Rapport d'essai - Essai de stérilité par la méthode de filtration sur membrane
Test report - Membrane filtration sterility testing

selon la Pharmacopée Européenne 9^e édition chapitre 2.6.1
according to the 9th European Pharmacopoeia Edition, chapter 2.6.1

ECHANTILLON(S)/SAMPLE(S)

Désignation : Product name :	Clearklens Bi-spore	Numéro de commande : Order number :	CF190368
Référence client : Customer reference :	100852440	POE : SIP :	NA
Numéro de lot : Batch number :	ENT41229 19 080	Matériau(x) : Material :	-
Date de réception : Receipt date :	mercredi 3 avril 2019	Donnée de stérilisation : Sterilization data :	-
Date de réalisation : Testing date :	jeudi 11 avril 2019	Commentaire(s) : Comment (s) :	10 bidons poolés
Nombre d'échantillon(s) : Sample quantity :	1		

PROTOCOLE/PROTOCOL

Volume d'échantillon : Sample volume :	50 ml	3 rinçages de 100 ml de DNP Thio	
Solution neutralisante : Neutralizing solution :	DNP + Thiosulfate	Volume de rinçage : Rinsing volume :	3 x 100 ml
Nombre de milieux testés : Number of Environment tested :	2	Volume d'immersion de la membrane : Immersion volume of membranes :	100 ml

Conditions Conditions	Milieux de culture Environnement of culture	Température d'incubation Incubation temperature	Durée d'incubation Incubation period
Bactéries aérobies, levures, champignons Aerobic and fungal	Bouillon Trypcase Soja Tryptane soy solution	22,5 ± 2,5°C	14 jours 14 days
Bactéries anaérobies et aérobies Anaerobic and aerobic	Bouillon Thioglycolate avec Résazurine Thioglycolate Resazurine	32,5 ± 2,5°C	14 jours 14 days

Validation de la méthode Method validation	09/OI/VAL.STE/017a
---	--------------------

RESULTATS / RESULTS

Examen de la croissance microbienne du milieu Examination of the microbial growth in the media			
Conditions / Milieux de culture Conditions / Media of culture	Après 7 jours After 7 days	Après 14 jours After 14 days	
Bactéries aérobies, levures et moisissures / Bouillon Trypcase Soja Aerobic and fungal / Tryptone soy solution	- Positifs Positive	0 Positifs Positive	
	- Négatifs Negative	1 Négatifs Negative	
Bactéries anaérobies et aérobies / Bouillon Thioglycolate avec Résazurine Anaerobic and aerobic / Thioglycolate Resazurine solution	- Positifs Positive	0 Positifs Positive	
	- Négatifs Negative	1 Négatifs Negative	

CONTRÔLES / CONTROLS

	Avant (Before)	Après (After)		
Contrôle plan de travail (UFC) Work plan control (CFU)	0	0	Contrôle de gant (UFC) : Glove control (CFU) :	6
Air sous flux laminaire (UFC) Laminar flow air control (CFU)	0		Contrôle stérilité milieux : Media sterility control :	Conforme

CONCLUSION

Les échantillons testés ne présentent pas de croissance microbienne après 14 jours d'incubation

Tested products doesn't shown any microbial development after 14 days of incubation.

Aucun produit ne s'est révélé positif lors de cet essai.

No Product was positive during the test.

Rédigé par :

Written by :

Françoise HINGREZ

Technicienne Biologiste
Biologist technician

Date :

lundi 29 avril 2019

Approuvé par :

Approved by :

Christine PARET

Directrice Qualité
Quality Manager

Le présent rapport ne peut être reproduit totalement ou partiellement qu'avec l'approbation du laboratoire et ne concerne que les objets soumis à l'essai.

This report may not be reproduced in whole or part without the approval of the laboratory and only involves the tested products.

Global Health Company

Microbiological and physico-chemical analysis - process validation

EN ISO 13485 (2016)

5, chemin du Catupolan - 69120 Vaulx-en-Velin - France

Tel. : 33 (0)4 72 81 22 62 - Fax : 33 (0)4 72 81 22 72

www.medicalgroup.fr

SEULE LA VERSION PAPIER SIGNÉE FAIT FOI
ONLY THE SIGNED PAPER VERSION IS AUTHENTIC

1/1
ENR75-TLBB-002B - Resultats STE PhEUR FM