



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 27.09.2019

Product Name	CLEARKLENS DE VH29
Product Code	7575783
Batch Number	FHP19203, 47639
Production Date	22/07/2019
Expiration Date	EXP 22/07/2021

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	Harabinda Panyas
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of Issuing : November 24 th 2017
--------------	--------------	--

R55480102

 09/13/2019 9:24:49 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2511290
P.O.#	PO33833			Sales Order #	2318458
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	09/12/2019 14:29:27 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	278	BX	101100418	Diversey Clearklens IPA/DE 900	FMP191911	WO47716 / DEV002080
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	703	BX	7515783	Diversey Clearklens DE 900 mL	FMP19203	WO47639 / DEV001333
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	369	BX	7513400	Diversey Clearklens IPA 900 mL	FMP19213	WO47635 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	5	BX	7515783 Sample	Diversey Clearklens DE 900 mL	FMP19203	WO47639 / DEV001333
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
105.000	5	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP19213	WO47635 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	1360	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	29.2 KGY	Pass	ESILLS	09/13/2019 07:26:38 GMT
		Reason Code Test				EMMA SILLS	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	42.8 KGY	Pass	ESILLS	09/13/2019 07:36:31 GMT
		Reason Code Test				EMMA SILLS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 09/13/2019 07:46:53 GMT

PHARMACEUTICAL ANALYSIS REPORT

Flexible Medical Packaging
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens DE

Sample Description: Sample Code: FMP19203 47639

Lucideon Sample Number: (195216)-40370

Lucideon Report Number: (195216)-40370/MFEP **Issue Number:** 1

Date Logged: 20-Sep-2019 **Order Number:** PO33926

Date Reported: 09-Oct-2019 **Date(s) of Test(s):** 23-Sep-2019 to 07-Oct-2019

Sterility Testing
Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



09 OCT 19

Mrs Andrea Saunders
Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited Registered in England No 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Kiens DE

Sample Description: Sample Code: FMP19203 47639

Lucideon Sample Number: (195216)-40371

Lucideon Report Number: (195216)-40371/ETEP Issue Number: 1

Date Logged: 20-Sep-2019 Order Number: PO33926

Date Reported: 30-Sep-2019 Date(s) of Test(s): 22-Sep-2019 to 22-Sep-2019

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

NoBoat 30-Sep-19

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1980455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com