



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 09.10.2019

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	FMP19232.47809
Production Date	20108/2019
Expiration Date	EXP 20108/2021

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.897

On behalf of Diversey site Quality Manager	Name :	Harokunda Pentyesb
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of Issuing : November 24 th 2017
--------------	--------------	--

R55480102

09/30/2019 14:47:09 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2520641
P.O.#	P033957			Sales Order #	2326874
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	09/28/2019 09:56:31 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	309	BX	7513400	Diversey Clearkens IPA 900 mL	FMP19235	WO47911 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	546	BX	7513400	Diversey Clearkens IPA 900 mL	FMP19232	WO47809 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	5	BX	7513400_SAMPLE	Diversey Clearkens IPA 900 mL	FMP19232	WO47809 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	880	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.7 KGY	Pass	PSAXTON	09/29/2019 00:15:26 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.5 KGY	Pass	PSAXTON	09/29/2019 00:15:45 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **ASHLEY HOBSON**
 Reason: **Work Order Completions**

Date: **09/30/2019 14:45:30 GMT**

R55480102

 10/03/2019 12:27:41 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2525084
P.O.#	PO33969			Sales Order #	2334066
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	10/02/2019 10:17:04 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	879	BX	7513400	Diversey Clearkens IPA 900 mL	FMP19232	WO47809 / DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	536	BX	7514372	Diversey Clearkens IPA 900 mL	FMP19241	WO48026 / DEV001250
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	5	BX	7514372 Sample	Diversey Clearkens IPA 900 mL	FMP19241	WO48026/ DEV001250
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	1420	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.6 KGY	Pass	TKVARDA	10/03/2019 10:58:23 GMT
		Reason Code Test				TOMAS KVARDA	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.2 KGY	Pass	TKVARDA	10/03/2019 10:58:49 GMT
		Reason Code Test				TOMAS KVARDA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 10/03/2019 12:26:54 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: FMP19232 47809

Lucideon Sample Number: (195535)-42390

Lucideon Report Number: (195535)-42390/MFEP Issue Number: 1

Date Logged: 07-Oct-2019 Order Number: 34035

Date Reported: 25-Oct-2019 Date(s) of Test(s): 09-Oct-2019 to 23-Oct-2019

Sterility Testing
Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Noted 25-Oct-19

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1860455.

Lucideon Limited
Queens Road, Pankhuff
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Kiens IPA

Sample Description: Sample Code: FMP19232 47809

Lucideon Sample Number: (195535)-42391

Lucideon Report Number: (195535)-42391/ETEP **Issue Number:** 1

Date Logged: 07-Oct-2019 **Order Number:** 34035

Date Reported: 23-Oct-2019 **Date(s) of Test(s):** 18-Oct-2019 to 22-Oct-2019

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Up Boot 23-Oct-19

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com