



Diversey Europe Operations BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427  
www.Diversey.com

## CERTIFICATE OF ANALYSIS

Date: 16/10/2019

Product Name	CLEARKLENS DE VH29
Product Code	7515983
Batch Number	FMP 19245 48035
Production Date	02/09/2019
Expiration Date	Exp 02/09/2021

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.880	0.888	0.881

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

*This document being issued electronically does not bear a signature*

COA Template	Version : 02	Date of issuing : November 24 <sup>th</sup> 2017
--------------	--------------	--

R55480102

 10/08/2019 7:38:06 GMT  
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2526683
P.O.#	PO34021			Sales Order #	2335837
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	10/05/2019 22:33:10 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	707	BX	7515783	Diversey Clearklens DE 900 mL	FMP19245	WO48036 / DEV001333
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	5	BX	7515783 Sample	Diversey Clearklens DE 900 mL	FMP19245	WO48036 / DEV001333
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	648	BX	7513400	Diversey Clearklens IPA 900 mL	FMP19254	WO48034/ DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	1360	BX	Total			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.8 KGY	Pass	PSAXTON	10/07/2019 09:51:28 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.6 KGY	Pass	PSAXTON	10/07/2019 09:51:44 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

 Electronically Signed By: ASHLEY HOBSON  
 Reason: Work Order Completions

Date: 10/08/2019 07:36:24 GMT

# PHARMACEUTICAL ANALYSIS REPORT

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens DE

**Sample Description:** Sample Code: Fmp19245 48036

**Lucideon Sample Number:** (195849)-44543

**Lucideon Report Number:** (195849)-44543/ETEP **Issue Number:** 1

**Date Logged:** 23-Oct-2019 **Order Number:** 34113

**Date Reported:** 05-Nov-2019 **Date(s) of Test(s):** 01-Nov-2019 to 01-Nov-2019

### Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


### Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

**End of Test Report**



05 NOV 19

Mrs Andrea Saunders

**Business Support Assistant**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com

# PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens DE

**Sample Description:** Sample Code: Fmp19245 48036

**Lucideon Sample Number:** (195849)-44542

**Lucideon Report Number:** (195849)-44542/MFEP

**Issue Number:** 1

**Date Logged:** 23-Oct-2019

**Order Number:** 34113

**Date Reported:** 15-Nov-2019

**Date(s) of Test(s):** 31-Oct-2019 to 14-Nov-2019

### Sterility Testing

Membrane Filtration EP


## Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report**



Mrs Andrea Saunders

**Business Support Assistant**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com