



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 04 March 2020

Product Name	CLEARKLENS IPA VH1	
Product Code	101100418	
Batch Number	FMP20041	49184
Production Date	10/02/2020	
Expiration Date	EXP 10/02/2022	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manger	Name:	Urszula Haraburda Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

02/24/2020 11:34:42 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2612006
P.O.#	PO34782			Sales Order #	2436024
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	02/21/2020 17:37:32 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	710	BX	101100418	Diversey Clearklens IPA/DE 900	FMP20041	WO49184/DEV002080
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	101100418_SAMPLE	Diversey Clearklens IPA/DE 900	FMP20041	WO49184/DEV002080
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	2	BX	101100418	Diversey Clearklens IPA/DE 900	TRIAL	T020/026
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	165	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20044	WO49156/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	880	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.9 KGY	Pass	TKVARDA	02/22/2020 05:29:09 GMT
		Reason Code Test				TOMAS KVARDA	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.4 KGY	Pass	TKVARDA	02/22/2020 05:29:31 GMT
		Reason Code Test				TOMAS KVARDA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
Reason: Work Order Completions

Date: 02/24/2020 11:33:03 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20041 49184

Lucideon Sample Number: UK201083-8333

Lucideon Report Number: UK201083-8333/MFEP **Issue Number:** 1

Date Logged: 05-Mar-2020 **Order Number:** 34891

Date Reported: 01-Apr-2020 **Date(s) of Test(s):** 18-Mar-2020 to 01-Apr-2020

Sterility Testing

Membrane Filtration EP


Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report


01 APR 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20041 49184

Lucideon Sample Number: UK201083-8334

Lucideon Report Number: UK201083-8334/ETEP **Issue Number:** 1

Date Logged: 05-Mar-2020 **Order Number:** 34891

Date Reported: 16-Mar-2020 **Date(s) of Test(s):** 11-Mar-2020 to 11-Mar-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

NcBook 16-Mar-20

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com