



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 16 April 2020

Product Name	CLEARKLENS DE VH29	
Product Code	100862174	
Batch Number	FMP20086	49158
Production Date	26/03/2020	
Expiration Date	EXP 26/03/2022	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.880	0.888	0.881

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

04/14/2020 9:08:49 GMT

Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2643046
P.O.#	PO35164			Sales Order #	2473389
	25 - 45 kGy	Gamma Irradiation	Irradiation Date/Time:	04/08/2020 16:02:49 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	12	BX	101101081	Diversey Clearklens DE 900ml	FMP20083	WO49600/DEV002096
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	8	BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP20083	WO49600/DEV002096
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	697	BX	100862174	Diversey Clearklens DE 900 mL	FMP20086	WO49158/DEV001910
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	3	BX	100862174 Sample	Diversey Clearklens DE 900 mL	FMP20086	WO49158/DEV001910
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	720	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.5 KGY	Pass	DMATHIAS	04/09/2020 09:45:40 GMT
		Reason Code Test				DAVID MATHIAS	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	41.8 KGY	Pass	DMATHIAS	04/09/2020 09:46:07 GMT
		Reason Code Test				DAVID MATHIAS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON
 Reason: Work Order Completions

Date: 04/14/2020 09:07:06 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens DE

Sample Description: Sample Code: Fmp20086 49158

Lucideon Sample Number: UK201736-13203

Lucideon Report Number: UK201736-13203/MFEP **Issue Number:** 1

Date Logged: 17-Apr-2020 **Order Number:** PO 35249

Date Reported: 05-May-2020 **Date(s) of Test(s):** 20-Apr-2020 to 04-May-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

N. Boot 05-May-20

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens DE

Sample Description: Sample Code: Fmp20086 49158

Lucideon Sample Number: UK201736-13204

Lucideon Report Number: UK201736-13204/ETEP **Issue Number:** 1

Date Logged: 17-Apr-2020 **Order Number:** PO 35249

Date Reported: 12-May-2020 **Date(s) of Test(s):** 09-May-2020 to 09-May-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



12 May 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com