

Diversey Europe Operations BV Maarasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

04 August 2020

Product Name	CLEARKLENS IPA VH1				
Product Code	7513400				
Batch Number	FMP20184	49641			
Production Date	02/07/2020				
Expiration Date	EXP 02/07/2022				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results	
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.873	

On behalf of Diversey site	Name:	Angelika Partynska urszula Haraburda
Quality Manger	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 07/30/2020 7:48:38 GMT

Page 1 of 1

Cianad Die

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2700911 P.O.# PO35902 Sales Order # 2543073

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 07/29/2020 08:32:36 GMT

SO Line #	Qty UO	M Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	595 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20184	WO49641/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20184	WO49641/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	322 BX	100862174	Diversey Clearklens DE 900 mL	FMP20189	WO49730/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	920 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.9 KGY	Pass	AWAGSTAFF	07/30/2020 06:36:20 GMT
		Reason Code Test				ADDISON WAGSTAFF	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.9 KGY	Pass	AWAGSTAFF	07/30/2020 06:36:41 GMT
		Reason Code Test				ADDISON WAGSTAFF	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 07/30/2020 07:46:31 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 07/30/

07/30/2020 7:51:25 GMT Page 1 of 1

Cianad Die

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2699238 P.O.# PO35875 Sales Order # 2541288

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 07/28/2020 23:43:46 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	122 BX	101100418	Diversey Clearklens IPA/DE 900	FMP20182	WO49646/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	101100418_SAMPLE	Diversey Clearklens IPA/DE 900	FMP20182	WO49646/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	835 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20184	WO49641/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	960 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	28.0 KGY	Pass	AWAGSTAFF ADDISON WAGSTAFF	07/30/2020 06:39:15 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	40.4 KGY	Pass	ESILLS EMMA SILLS	07/30/2020 07:21:04 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 07/30/2020 07:49:18 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20184 49641

Lucideon Sample Number: UK203165-22402

Lucideon Report Number: UK203165-22402/MFEP Issue Number: 1

Date Logged: 05-Aug-2020 Order Number: PO36053

Date Reported: 04-Sep-2020 Date(s) of Test(s): 20-Aug-2020 to 03-Sep-2020

Sterility Testing
Membrane Filtration EP

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

End of Test Report

NOBOOF OUSEP-20

Mrs Natalie Boot

Senior Business Support Assistant

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20184 49641

Lucideon Sample Number: UK203165-22403

Lucideon Report Number: UK203165-22403/ETEP

Issue Number:

Date Logged:

05-Aug-2020

Order Number:

PO36053

Date Reported:

04-Sep-2020

Date(s) of Test(s):

18-Aug-2020 to 27-Aug-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

NORME OU SEP-20

Mrs Natalie Boot

Senior Business Support Assistant

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation