



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 08 February 2021

Product Name	CLEARKLENS IPA VH1		
Product Code	7513400		
Batch Number	FMP21012	51287	
Production Date	12/01/2021		
Expiration Date	EXP 12/01/2023		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits Lower - Upper		Results
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.875

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Justyna Staron
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

 02/01/2021 10:11:33 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2814855
P.O.#	PO37250			Sales Order #	2682480
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	01/30/2021 23:27:07 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	465	BX	7513400	Diversey Clearklens IPA 900 mL	FMP21012	WO51287/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	15	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21012	WO51287/DEV00152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	480	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	PSAXTON	01/31/2021 18:24:59 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.8 KGY	Pass	PSAXTON	01/31/2021 18:25:16 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 02/01/2021 10:08:11 GMT

R55480102

 02/01/2021 10:16:54 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2817680
P.O.#	PO37249			Sales Order #	2685609
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	01/30/2021 19:56:34 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	960	BX	7513400	Diversey Clearklens IPA 900 mL	FMP21012	WO51287/DEV001152
			Dose Map	302_0030		
	960	BX	Total	Ref Pos: 10E1, AF Min = 0.71		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.8 KGY	Pass	PSAXTON	01/31/2021 18:22:12 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.3 KGY	Pass	PSAXTON	01/31/2021 18:22:39 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 02/01/2021 10:13:36 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp21012 51287

Lucideon Sample Number: UK21568-3973

Lucideon Report Number: UK21568-3973/ETEP

Issue Number: 1

Date Logged: 09-Feb-2021

Order Number: PO 37402

Date Reported: 25-Feb-2021

Date(s) of Test(s): 24-Feb-2021 to 24-Feb-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

NB Boot 25-Feb-21

Mrs Natalie Boot

Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp21012 51287

Lucideon Sample Number: UK21568-3972

Lucideon Report Number: UK21568-3972/MFEP

Issue Number: 1

Date Logged: 09-Feb-2021

Order Number: PO 37402

Date Reported: 26-Feb-2021

Date(s) of Test(s): 11-Feb-2021 to 25-Feb-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

no Book 26-feb-21

Mrs Natalie Boot

Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com