

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

02 July 2021

Product Name	CLEARKLENS DE VH29				
Product Code	101101081				
Batch Number	FMP21146	51895			
Production Date	26/05/2021				
Expiration Date	EXP 26/05/2023				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results Clear Colourless Liquid	
Appearance	Visual	Clear Colou	rless Liquid		
Specific Gravity (20°C)	DM004	0.880	0.888	0.883	

On behalf of Diversey site	Name:	Justyna Staron Ewelina Marki	
Quality Manger	Position	Quality	Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

06/21/2021 12:28:57 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2906966 P.O.# 2795675

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 06/19/2021 01:32:35 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	878 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21140	WO51841/STR001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21140	WO51841/STR001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
103.000	69 BX	101101081	Diversey Clearklens DE 900ml	FMP21146	WO51895/STR002096
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
	950 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.6 KGY	Pass	JBEARNE	06/19/2021 16:34:06 GMT
		Reason Code Test				JARETH BEARNE	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	41.0 KGY	Pass	JBEARNE	06/19/2021 16:34:26 GMT
		Reason Code Test				JARETH BEARNE	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/21/2021 12:28:39 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 06/21/2021 12:21:25 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2906979 P.O.# 400038 Sales Order # 2795682

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 06/19/2021 05:31:13 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	211 BX	100862174	Diversey Clearklens DE 900 mL	FMP21148	WO51896/STR001910
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
102.000	636 BX	101101081	Diversey Clearklens DE 900ml	FMP21146	WO51895/STR002096
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
103.000	3 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP21146	WO51895/STR002096
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
	850 BX	Total			

Quality Test Summary

				or ourman,	,	Signed By	-
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.5 KGY	Pass	JBEARNE JARETH BEARNE	06/19/2021 16:47:46 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.5 KGY	Pass	JBEARNE JARETH BEARNE	06/19/2021 16:48:04 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/21/2021 12:21:07 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens DE

Sample Description:

Sample Code: Fmp21146 51895

Lucideon Sample Number: UK212880-19589

Lucideon Report Number: UK212880-19589/ETEP

Issue Number:

Date Logged:

05-Jul-2021

Order Number:

PO 400228

Date Reported:

Date(s) of Test(s):

06-Jul-2021

to 06-Jul-2021

12-Jul-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Mrs Rachel McGrandle

Pharmaceutical Microbiology Manager

Page 1 of 1

Stoke-on-Trent

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens DE

Sample Description:

Sample Code: Fmp21146 51895

Lucideon Sample Number: UK212880-19588

Lucideon Report Number: UK212880-19588/MFEP

Issue Number:

Date Logged:

05-Jul-2021

Order Number:

PO 400228

Date Reported:

23-Jul-2021

Date(s) of Test(s):

07-Jul-2021

to 21-Jul-2021

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Mrs Andrea Saunders

Business Support Administrator