

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

20 May 2021

Product Name	CLEARKLENS DE VH29				
Product Code	100862174				
Batch Number	FMP21110	51686			
Production Date	20/04/2021				
Expiration Date	EXP 20/04/2023				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method Limits Lower - Upper			Results	
Appearance	Visual	Clear Color	ırless Liquid	Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.880	0.888	0.887	

On behalf of Diversey site	Name:	Ewelina Markiel urszula Haraburo		
Quality Manger	Position	Quality	Control Inspector	

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

05/10/2021 10:11:20 GMT Page 1 of 1

Cianad Die

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2873070 P.O.# PO37898 Sales Order # 2754220

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 05/08/2021 03:26:54 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	457 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21091	WO51421/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21091	WO51421/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	440 BX	100862174	Diversey Clearklens DE 900 mL	FMP21110	WO51686/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	900 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.5 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:48:45 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.7 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:49:06 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 05/10/2021 10:11:01 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 05

05/10/2021 9:56:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2876905 P.O.# PO37942 Sales Order # 2759058

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 05/07/2021 22:31:30 GMT

SO Line #	Qty UOI	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	265 BX	100862174	Diversey Clearklens DE 900 mL	FMP21110	WO51686/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	100862174 Sample	Diversey Clearklens DE 900 mL	FMP21110	WO51686/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	592 BX	7514372	Diversey Clearklens IPA 900 mL	FMP21111	WO51681/DEV001250
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	860 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.6 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:50:16 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.4 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:50:35 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 05/10/2021 09:56:30 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens DE

Sample Description:

Sample Code: FMP21110 51686

Lucideon Sample Number: UK212191-15215

Lucideon Report Number: UK212191-15215/MFEP

Issue Number:

Date Logged:

20-May-2021

Order Number:

PO 38115

Date Reported:

11-Jun-2021

Date(s) of Test(s):

25-May-2021 to 08-Jun-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

JC800111-JUN-21

Mrs Natalie Boot

Business Support Administrator

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens DE

Sample Description:

Sample Code: FMP21110 51686

Lucideon Sample Number: UK212191-15216

Lucideon Report Number: UK212191-15216/ETEP

Issue Number:

Date Logged:

20-May-2021

Order Number:

PO 38115

Date Reported:

01-Jun-2021

Date(s) of Test(s):

21-May-2021 to 26-May-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000EU/ml)

Tests carried out in accordance with cGMP Refer to laboratory investigation report (LIR) 0698.

End of Test Report



Mrs Andrea Saunders

Business Support Administrator

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation