

Diversey Europe Operations BV Magrasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

13 October 2022

Product Name	CLEARKLENS DE VH29					
Product Code	101101081					
Batch Number	FMP22265	55569				
Production Date	22/09/2022					
Expiration Date	EXP 22/09/2024					

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi	its	Results		
		Lower	- Upper			
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid		
Specific Gravity (20°C)	DM004	0.880	0.888	0.883		

On behalf of Diversey site Quality Manger	Name:	Partyrisher, MStanon
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th 2017



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 1

10/03/2022 10:44:24 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 3203698 P.O.# Sales Order # 3143157

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/01/2022 18:31:39 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	325 BX	101101081	Diversey Clearklens DE 900ml	FMP22265	55569/STR002096
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	24 BX	101105062	Diversey Clearklens IPA 900ml	FMP22228	55367/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	1 BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22228	55367/DEV002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
	350 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	- Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.8 KGY	Pass	IWEATHERAL IAN WEATHERALL	10/02/2022 12:38:56 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.5 KGY	Pass	IWEATHERAL IAN WEATHERALL	10/02/2022 12:39:18 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: CLAIRE CICALA Date: 10/03/2022 10:40:12 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 10/1

10/10/2022 9:28:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 3207097 P.O.# 403522 Sales Order # 3147342

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/09/2022 03:15:52 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	384 BX	101101081	Diversey Clearklens DE 900ml	FMP22265	55569/STR002096
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP22265	55569/STR002096
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	538 BX	101105062	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
	925 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.7 KGY	Pass	PSAXTON	10/09/2022 16:45:26 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.1 KGY	Pass	PSAXTON	10/09/2022 16:45:50 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: CLAIRE CICALA Date: 10/10/2022 09:28:21 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster

LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens DE

Sample Description:

Sample Code: Fmp22265 55569

Lucideon Sample Number: UK224033-26352

Lucideon Report Number: UK224033-26352/MFEP

Issue Number:

Date Logged:

14-Oct-2022

Order Number:

403650

Date Reported:

04-Nov-2022

Date(s) of Test(s):

20-Oct-2022 to 03-Nov-2022

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Miss Rosie Jones

Pharmaceutical Microbiology Manager

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens DF

Sample Description:

Sample Code: Fmp22265 55569

Lucideon Sample Number: UK224033-26353

Lucideon Report Number: UK224033-26353/ETEP

Issue Number:

Date Logged:

14-Oct-2022

Order Number:

403650

Date Reported:

Date(s) of Test(s):

18-Oct-2022 to 18-Oct-2022

25-Oct-2022

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Robin Williams

Microbiology Laboratory Supervisor

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation