

Diversey Europe BV
Maarssenbroeksedijk 2
3542 DN Utrecht, The Netherlands
Tel. +31 (0)30 247 6427
www.Diversey.com
Diversey - Lets Live

CERTIFICATE OF ANALYSIS

Date issue: 04 November 2022

Product Name	CLEARKLENS IPA VH1			
Product Code	☑ 101105062	□ 101105063	□ 101	105064
Batch Number	FMP22270		55570	
Production Date	27/09/2022			
Expiration Date	EXP 27/09/2024			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product

Test	Test Method	Speci	fication	UOM	Results
Appearance	Visual	Clear Colour	less Liquid	N/A	Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	g/cm3	0.875

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference No:
Sterilization	Gamma Irradiation	25.0 - 45.0 (kGy)	YES	55570
Sterility Certificate	Sterility by Mebrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	YES	UK224304- 28004
2.6.14 Bacterial Endotoxins	Method C - Turbidimetric Kinetic Method (EP/USP Criteria)	<0.2 EU/mL	YES	UK224304- 28005



The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommeded conditions indicated in the SDS.

End of certificate of analysis



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 10/1

10/10/2022 9:28:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 3207097 P.O.# 403522 Sales Order # 3147342

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/09/2022 03:15:52 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	384 BX	101101081	Diversey Clearklens DE 900ml	FMP22265	55569/STR002096
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP22265	55569/STR002096
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	538 BX	101105062	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
	925 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	26.7 KGY	Pass	PSAXTON PAUL SAXTON	10/09/2022 16:45:26 GMT
450.00	Maximum Dose	25.0 kGv	45.0 kGv	38.1 KGY	Pass	PSAXTON PSAXTON	10/09/2022 16:45:50 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: CLAIRE CICALA Date: 10/10/2022 09:28:21 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

10/10/2022 9:28:55 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 3207897 P.O.# 403541 Sales Order # 3148407

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/09/2022 06:53:51 GMT

so UOM **Customer Item Number Customer Item Description** Customer Qty Customer Line # Lot Number Load Number 101.000 675 ВХ 101105062 Diversey Clearklens IPA 900ml FMP22270 55570/STR002199 Dose Map 302 0112 RF: 13.5A3, AF Min: 0.717 675 BX Total

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.5 KGY	Pass	PSAXTON	10/09/2022 16:43:10 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.3 KGY	Pass	PSAXTON	10/09/2022 16:43:31 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: CLAIRE CICALA Date: 10/10/2022 09:27:55 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 10/

10/31/2022 11:46:29 GMT Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 3218848 P.O.# 33161321

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/28/2022 18:57:31 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	225 BX	101105062	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3 BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	722 BX	101105062	Diversey Clearklens IPA 900ml	FMP22277	55571/STR002199
	Dose Map	302_0112	RF: 13.5A3, AF Min: 0.717		
	950 BX	Total			

Quality Test Summary

			Quality 10	or ourman	,	Signed By	-
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.7 KGY	Pass	DMATHIAS DAVID MATHIAS	10/29/2022 04:59:11 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.6 KGY	Pass	ESILLS EMMA SILLS	10/31/2022 09:43:53 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/31/2022 11:38:55 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens IPA

Sample Description:

Sample Code: Fmp22270 55570

Lucideon Sample Number: UK224304-28004

Lucideon Report Number: UK224304-28004/MFEP

Issue Number:

Date Logged:

04-Nov-2022

Order Number:

403875

Date Reported:

25-Nov-2022

Date(s) of Test(s):

10-Nov-2022 to 24-Nov-2022

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

75 mw 22 Miss Rosie Jones

Pharmaceutical Microbiology Manager

Page 1 of 1

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

ClearKlens IPA

Sample Description:

Sample Code: Fmp22270 55570

Lucideon Sample Number: UK224304-28005

Lucideon Report Number: UK224304-28005/ETEP

Issue Number:

Date Logged:

04-Nov-2022

Order Number:

403875

Date Reported:

30-Nov-2022

Date(s) of Test(s):

12-Nov-2022 to 12-Nov-2022

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

30MV27

Miss Rosie Jones

Pharmaceutical Microbiology Manager

Page 1 of 1