

CERTIFICATE OF ANALYSIS

Date issue: 11 November 2022


Product Name	CLEARLENs IPA VH1		
Product Code	<input checked="" type="checkbox"/> 101105062	<input type="checkbox"/> 101105063	<input type="checkbox"/> 101105064
Batch Number	FMP22277	55571	
Production Date	04/10/2022		
Expiration Date	EXP 04/10/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product

Test	Test Method	Specification	UOM	Results
Appearance	Visual	Clear Colourless Liquid	NA	Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872 0.883	g/cm3	0.875

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference No:
Sterilization	Gamma Irradiation	25.0 - 45.0 (kGy)	YES	55571
Sterility Certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	YES	UK224456 - 28884
2.6.14 Bacterial Endotoxins	Method C - Turbidimetric Kinetic Method (EP/USP Criteria)	<0.2 EU/mL	YES	UK224457 - 28890

	Date of released on behalf of Diversey's Manufacturing Location:	Testers:	
		Justyna Staron	11/11/2022
		Sabcho Svetoslavov	11/11/2022
		Quality Control Inspector	
		Olga Kirchner	05/12/2022

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

End of certificate of analysis

R55480102

10/31/2022 11:46:29 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3218848
P.O.#	403697			Sales Order #	3161321
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	10/28/2022 18:57:31 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	225	BX	101105062	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22270	55570/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	722	BX	101105062	Diversey Clearklens IPA 900ml	FMP22277	55571/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
	950	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.7 KGY	Pass	DMATHIAS	10/29/2022 04:59:11 GMT
		Reason Code Test				DAVID MATHIAS	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.6 KGY	Pass	ESILLS	10/31/2022 09:43:53 GMT
		Reason Code Test				EMMA SILLS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
Reason: Work Order Completions

Date: 10/31/2022 11:38:55 GMT

R55480102

 11/07/2022 8:40:09 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3227106
P.O.#	403801			Sales Order #	3171125
	25 - 45 kGy	Gamma Irradiation		Irradiation Date/Time:	11/05/2022 17:25:56 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	709	BX	101105062	Diversey Clearklens IPA 900ml	FMP22277	55571/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22277	55571/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	238	BX	101105063	Clearklens IPA 900 mls	FMP22297	55946/STR002201
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
	950	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.8 KGY	Pass	PHAMPSON	11/06/2022 05:38:25 GMT
		Reason Code Test				PIERS HAMPSON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.6 KGY	Pass	PHAMPSON	11/06/2022 05:38:46 GMT
		Reason Code Test				PIERS HAMPSON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 11/07/2022 08:39:57 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: Fmp22277 55571

Lucideon Sample Number: UK224456-28884

Lucideon Report Number: UK224456-28884/MFEP **Issue Number:** 1

Date Logged: 14-Nov-2022 **Order Number:** 403932

Date Reported: 02-Dec-2022 **Date(s) of Test(s):** 17-Nov-2022 to 01-Dec-2022

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Natalie Boot 02 Dec 22

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: Fmp22277 55571

Lucideon Sample Number: UK224457-28890

Lucideon Report Number: UK224457-28890/ETEP **Issue Number:** 1

Date Logged: 14-Nov-2022 **Order Number:** 403932

Date Reported: 30-Nov-2022 **Date(s) of Test(s):** 20-Nov-2022 to 20-Nov-2022

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



Miss Rosie Jones

Pharmaceutical Microbiology Manager

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com