

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

### **CERTIFICATE OF ANALYSIS**

Date: 05 July 2022

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
<b>Product Code</b>	7515828		
Batch Number	FMP22186 54510		
Production Date	05/07/2022		
Expiration Date	EXP 05/07/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	12.0
Specific Gravity (20°C)	DM004	1.004	1.020	1.005
Potential Chlorine Dioxide		100	180	121.5

On behalf of Diversey site	Name:	Justyna Staron - Angelika Partynska
Quality Manger	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version : 02 Date of issuing : November 24th 2017



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

## **CERTIFICATE OF ANALYSIS**

Date: 05 July 2022

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26 first off		
<b>Product Code</b>	7515828		
Batch Number	FMP22186 54510		
<b>Production Date</b>	05/07/2022		
<b>Expiration Date</b>	EXP 05/07/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limit	ts	Results
		Lower	- Upper	
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	1.9
Specific Gravity (20°C)	DM004	1.010	1.030	1.015

On behalf of Diversey site	Name:	Justyna Staron Angelika Partynska
Quality Manger	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing: November 24th 2017
CONTRIBITION	V C/3/C/1 : 02	Baic of issuing . No terriber 2 mil 2017

## **STERIS: Gamma Certificate Of Processing**

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-27637A

Product CodeLot NumberQuantityUOMBI-SP B3 Components DV5857545158Case

Validation Reference Number: 5857

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 30.9 Reference Dose Range (kGy): 26.1 - 35.8 Calculated Max Dose (kGy): 37.7

PO Number: 402391

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### Signature Manifest

Reviewed and E-Signed By: Nelia Dias (Quality Engineer)

Date/Time E-Signed: 19-Apr-2022 9:46 AM

Document Content Revision: 1

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

#### **Processing Location**

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO54515 will be used in finished batch Fmp22186 54510.

Olga Kirchner

Document ID: 109186

N/A Last Revised in Rel 2.0.0.0 Rel Date: 13-Aug-2018 Page 1 of 1

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp22186 54510

Lucideon Sample Number: UK222895-18561

Lucideon Report Number: UK222895-18561/MFEP

Issue Number:

1

**Date Logged:** 

18-Jul-2022

**Order Number:** 

403033

**Date Reported:** 

09-Aug-2022

Date(s) of Test(s):

20-Jul-2022

to 03-Aug-2022

**Sterility Testing** 

Membrane Filtration EP

**Test Results:** 

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 

25-10-A-RD 1008DL

Mrs Natalie Boot

Senior Business Support Administrator

# PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp22186 54510

Lucideon Sample Number: UK222895-18562

Lucideon Report Number: UK222895-18562/MFEP

Issue Number:

Date Logged:

18-Jul-2022

**Order Number:** 

403033

**Date Reported:** 

09-Aug-2022

Date(s) of Test(s):

20-Jul-2022

to 03-Aug-2022

**Sterility Testing** 

Membrane Filtration EP

**Test Results:** 

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 

BOOK 09-A9-22 Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.