



Diversey Europe Operations BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427  
www.Diversey.com

## CERTIFICATE OF ANALYSIS

Date 15 March 2023

Product Name	CLEARKLENS CLEANSINALD SC VH9		
Product Code	7516429		
Batch Number	FMP23073	56393	
Production Date	14/03/2023		
Expiration Date	EXP 14/03/2025		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
pH (neat solution)	DM001	12.0	13.0	12.9
Specific Gravity (20°C)	DM004	1.040	1.060	1.047
Cationic Content	DM020	14.25	15.75	14.81

On behalf of Diversey site Quality Manger	Name:	Edyta Rodrigues Justyna Staron
	Position	Quality Control Inspector

*This document being issued electronically does not bear a signature*

COA Template	Version : 04	Date of issuing : March 11th 2019
--------------	--------------	-----------------------------------

# STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-34942A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
C/SINALD BOTTLES/CAPS/BAGS DV4894	56442	16	Case

Validation Reference Number: 4894

Processing Run Start Date: 11-Feb-2023 10:34 PM

Processing Run End Date: 12-Feb-2023 4:10 AM

Specified Dose Range (kGy):	25.0 - 40.0	Calculated Min Dose (kGy):	26.7
Reference Dose Range (kGy):	27.4 - 36.5	Calculated Max Dose (kGy):	32.5

PO Number: 404534

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

**Gamma Process Run Approval authorized by STERIS**

Date/Time E-Signed: 12-Feb-2023 8:22 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

## Processing Location

Synergy Health Sterilisation UK  
Limited, a STERIS Company  
Brunel Close  
Drayton Fields Industrial Estate  
Daventry  
Northants  
NN11 8RB  
Phone: + 44(0) 1327 706 111

Items irradiated under WO56442 will be used in finished batch Fmp23073 56393.

Olga Kirchner

Document ID: 13543

N/A

Last Revised in Rel 2.0.0.0

Rel Date: 13-Aug-2018

Page 1 of

# PHARMACEUTICAL ANALYSIS REPORT



## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Cleansinald SC

**Sample Description:** Sample Code: Fmp23073 56393

**Lucideon Sample Number:** UK231004-6857

**Lucideon Report Number:** UK231004-6857/MFEP

**Issue Number:** 1

**Date Logged:** 17-Mar-2023

**Order Number:** 404844

**Date Reported:** 04-Apr-2023

**Date(s) of Test(s):** 20-Mar-2023 to 03-Apr-2023

### Sterility Testing

Membrane Filtration EP

## Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report**

Natalie Boot 04 APR 23

Mrs Natalie Boot

**Senior Business Support Administrator**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com