



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 16-05-2023

Product Name	ClearKlens IPA 3x 1000L + filt. LN7
Product Code	101101333
Batch Number	MUL 504777
Manufacturing Date	15-05-2023
Expired Date	15-05-2025

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

The Manufacturing/Expiration dates refer to packing date of the bulk product. As the packaging process could last for more than 1 day, a difference between the dates mentioned on this document and those printed on delivered stock is possible. For stock management purposes it is recommended that the dates printed on the packs are used.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance / Colour	MF01	Colorless Liquid		Conform
Odour	MF02	Alcohol		Conform
pH (Neat)	MF09 / DM-001	6,00 - 9,00		8,90
Specific Gravity (g/ml at 20°C)	MF08 / DM-004	0,872 - 0,883		0,875

Manufacturer: Multifill BV Constructieweg 25A 3641 SB Mijdrecht The Netherlands	Name :	S. Raave
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

OOS COA Template	Version : 04	Date of issuing : March 11 th 2019
------------------	--------------	---