



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 20-03-2026

Product Name	DI ClearKlens IPA VH1 2x5L W4502
Product Code	101105065
Batch Number	MUL 529710
Manufacturing Date	04-02-2026
Expired Date	04-02-2028

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

The Manufacturing/Expiration dates refer to packing date of the bulk product. As the packaging process could last for more than 1 day, a difference between the dates mentioned on this document and those printed on delivered stock is possible. For stock management purposes it is recommended that the dates printed on the packs are used.

Test	Test Method	Limits Lower - Upper	Results
Appearance / Colour	MF01	Clear Colorless Liquid	Conform
Density (g/ml at 20°C)	MF08 / DM-004	0,872 – 0,883	0,873
pH (Neat)	MF09 / DM-001	6,00 – 9,00	7,90

Manufacturer: Multifill BV Constructieweg 25A 3641 SB Mijdrecht The Netherlands	Name	M. Süer
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

OOS COA Template	Version : 04	Date of issuing : March 11 th 2019
------------------	--------------	---