

CERTIFICATE OF ANALYSIS

Date issue: 10-Mar-21

Product Name	ClearKlens Tego 2000SC VH25
Product Code	100868202
Batch Number	FMP20202 50113
Production date	20/07/2020
Expiration Date	20/07/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Conformity
Appearance	Visual	Liquid Clear Slightly Yellow	N/A	Liquid Clear Slightly Yellow
Specific Gravity (20°C)	DM 004	0.993 – 1.003	g/cm ³	1.002
pH (neat Solution)	DM 001	7.5 – 8.5	N/A	8.1

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 45.0 (kGy)	Conform	2173-12305A
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	Pass	UK203165-22406/MFEP

	Date of released On behalf of Diversey's Manufacturing location: 04.08.2020	by : Urzula Haraburda & Angelika Partynska Position: Quality Control
---	---	---

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

End of certificate of analysis