

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 16/7/12

Product Name	CLEARKLENS IPA VH1
Product Code	7514372
Batch Number	FMP12179 32304
Production Date	06/2017
Expiration Date	06/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits Lower - Upper	Results
Appearance	Visual	Clear Colourless Liquid	CIBOT COLOUPES
Specific Gravity (20°C)	JDM004	0.872 0.88	

On behalf of Diversey site Quality Manager	Name:	Chistoc
	Position	Quality Control Inspector

Inis document being issued el	ectronically does not bear a signature	
COA 7514372	Version: 03	Date of issuing: November 28th 2011
		1 - 100 of totaling . Ivo vehilber 20 2011

CERTIFICATE OF IRRADIATION Number: B

0538016 00

FLE001

Flexible Medical Packaging Ltd

Unit 8

White Cross Industria Estate

Hightown Lancaster

LA1 4XS

Cust. Ref: Date Rec'd Date:

F018787 5/07/12 9/07/12

Page 1

Synergy Health pic

Roydsdale Way · Euroway Trading Estate Bradford · West Yorkshire

BD4 6SE · United Kingdom Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061 Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

Date: 77017	16		
ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212070178	JD IPA BOTTLES 25-45kGy	715	B/No.W032304 - CUSTOMER B/No.7514372 FMP 12179 20 21612 ROUTINE DOSE: 40.5, 39.8, 36.4, 36.0kGy.
B2FLE0010999 B212070178	SAMPLES	5	
			CUSTOMER BATCH/CODE No.s AS ORDER No.PO18787 DATED 4/7/12
	Despatch Text:		
	14PLTS		
	Total	720	Last Page 1 of 1
		All in accor	rdance with current

All in accordance with current Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories Limited, Hoeford Point, Barwell Lane, Gosport Hampshire P013 0AU England

Telephone: +44(0)1329 226600 +44(0)1329 832511

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

16 Jul 2012

Date Tested:

18 Jul 2012

Date Test Completed:

18 Jul 2012

Purchase Order:

PO18835

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0005708/2

Test Required: Date Received: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

16/07/2012

Test Article: Sample Code: Clearklens IPA Block 2 FMP12179 32304

3592

Batch Ref: Qty Received:

1 x 900mL bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2
Spiked Recovery	MM110/01	92 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Shor Lund.

Date: 19 Jul 2012 15:50:02

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0005708

Certificate of Analysis - OSMM

Print Number: P0004184

Page 1 of 1



Wickham Laboratories Limited, Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600

+44(0)1329 832511

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS

Date Received:

16 Jul 2012

Date Tested:

20 Jul 2012

Date Test Completed:

03 Aug 2012

Purchase Order:

PO18835

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

Test Required:

Dale Received:

Test Article:

Sample Code:

Batch Ref: **Qty Received:** 0005708/1

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

16/07/2012

Clearkiens IPA Block 2 FMP12179 32304

3592

20 x 900mL bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03
Comments	i	

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Ehm Wood

Date: 07 Aug 2012 17:46:43

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Lini

Consignment: 0005708

Certificate of Analysis - OSMM

Print Number: P0004564

Page 1 of 1