



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 17/8/12

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP12220 32540
Production Date	08/2012
Expiration Date	06/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly yellow liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.5
Specific Gravity (20°C)	JDM004	0.980	1.020	0.997

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513403	Version : 04	Date of issuing : April 19 th 2012
-------------	--------------	---

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com



LA1 4XS

Cust. Ref: P018964

Date Rec'd 9/08/12

Date: 14/08/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B212080340	TEGO 2001 SS 25-45kGy	108	B/No. W032540 - CUST/BN0.FMP12220 ROUTINE DOSE 38.1, 36.2, 36.9 kGy
B2FLE0010093 B212080340	TEGO 2001 SS 25-45kGy	26	B/No. W032657 - CUST/BN0.FMP12221 ROUTINE DOSE 38.1, 36.2, 36.9 kGy
B2FLE0010999 B212080340	SAMPLES	6	1 X B/No. W032540 5 X B/No. W032657 CUSTOMER BATCH/CODE No.s AS ORDER No. P018964 DATED 9/8/12
Despatch Text: 3PLTS			
Total		140	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC