



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 26.09.12

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	7516431
Batch Number	FMP 12270 32303
Production Date	09/2012
Expiration Date	03/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		clear slightly yellow liquid
pH (100% 20°C)	JDM001	9.0	11.4	11.0
Specific Gravity (20°C)	JDM004	0.990	1.010	0.997

On behalf of Diversey site Quality Manager	Name :	Martens
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7516431	Version : 04	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@synergyhealthplc.com
Website: www.synergyhealthplc.com



LA1 4XS

P018919

Cust. Ref: 2/08/12

Date Rec'd 6/08/12

Date: 6/08/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010129 B212080121	C/SINALD SS 900ml BOTTLE 25-40kGy	11	B/No.W032302 - CUSTOMER B/No.32302 FMP12270 32303, (23/08/12) MAX....35.9....MIN....27.1....kGy
Despatch Text: 2PLTS			
Total		11	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *J. (Offic)*

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 28 Sep 2012
Date Tested: 04 Oct 2012
Date Test Completed: 18 Oct 2012
Purchase Order: PO19207

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0006911/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 28/09/2012
Test Article: Clearklens Clensinald SS
Sample Code: 2981
Batch Ref: FMP12270 32303
Qty Received: 20 x 900mL Bottles + 4 spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

S.L. Wood

Date: 22 Oct 2012 09:06:03

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0006911

Certificate of Analysis - OSMM

Print Number: P0006235

Page 1 of 1